

	State of Alaska Scientific Crime Detection Laboratory		
	DataMaster Calibration Form - DMCAF2011R1		
	Issuing Authority - Nita Bolz, Scientific Director		
Approval date: 03/10/11	Effective date: 03/10/11	Archive date: Active	

DataMaster DMT Calibration Form

Instrument # 100404

Date: 5/18/11

1. Wet Bath Simulator Solution

- a. Level 0.100
- b. Manufacturer Guth
- c. Lot 10260
- d. Expiration Date 10/05/12
COB 4/1/11

2. Wet Bath Simulator

- a. Serial Number DB 4161
- b. Date Calibrated 8/23/10

3. Attach Calibration Factors

Breath Alcohol Analyst C. O' Bryant

CALIBRATION FACTORS

Alaska Department of Public Safety
DAZMASTER dnt: 100404

Date: 05/18/2011
Time: 14:02:44

SUPERVISOR NAME:
COLLEEN S. O'BRYANT
CERTIFICATE #: 6237

LOT #: GUTH100
BOTTLE #: 10260
EXPIRATION: ~~01/01/2012~~ *05/19/11*
10/15/2011

Ca	=	0.1000		
CAL	=	0.970035	0.800	<= CAL < 1.200
b1	=	0.0085	0.0000	<= b1 < 0.0040
b2	=	0.0071	0.0010	<= b2 < 0.0100
b3	=	0.0001	0.0000	<= b3 < 0.0040
Xq	=	0.0888	0.0500	<= Xq < 0.2000
a21	=	1.183100	1.050	<= a21 < 1.300
a31	=	0.434619	0.300	<= a31 < 0.800

COB 2/2

Final page for report

\

\\ANCSCLSAN01\JTRAX\CASEDOCS\B-100404\B-100404_ECF-20140

2/18/2014