

In-field Instrument Review

Dates: 1/3/2015 – 2/21/2015

Review By: Colleen O'Bryant

100346 AST/Soldotna: (2) RFI, (2) Blank Error – No action taken
100347 AST/Tok: (1) Blank Error – No action taken
100350 Ketchikan PD: (3) Interference Detected on a subject test, all three same subject – No action taken
100354 AST/Fairbanks: (1) RFI, (1) Pump Error, (1) Blank Error – No action taken
100361 Unalaska DPS: (2) Standard Out of Range, (1) Ambient Fail – tank change performed following three status messages, no further issue – No action taken
100369 North Pole PD: (1) RFI – No action taken
100373 Seward PD: (1) RFI – No action taken
100376 Metlakatla PD : (1) Pump Error – No action taken
100378 AST/Talkeetna: (1) Ambient Fail – No action taken
100384 NSBPD: (1) Internal Standard Error – No other issues – No action taken
100392 Kodiak PD: (2) Standard Out of Range – no tank change – issue hasn't been reproduced – No action taken
100397 Ft. Greely: (1) RFI, (1) Blank Error – removed from service 1/10/15
100400 Palmer PD: (2) RFI, (1) Interference on External Standard, (1) Filter Wheel Error – Spoke with Luke Szipsky (Palmer PD) and are going to monitor the instrument
100402 Sitka PD: (2) Interference Detected on a subject test, (1) Internal Standard Error – Watching for more data – No action taken
100406 Kenai PD: (2) RFI – No action taken
100411 Kenai PD: (2) Tank Interference, (1) Blank Error – Removed from service 1/24/2015
100421 JBER: (2) Tank Interference, (1) Internal Standard Error, (1) Detector Overflow – All status messages during testing scenarios – No action taken
100422 JBER: (2) RFI, (1) Ambient Fail – No action taken
100424 Eielson: (1) Standard Out of Range
100666 APD: (12) Interference Detected (12 of 149 tests) – Removed from service 2/17/2015, instrument froze during software update.
100678 Wasilla PD: (1) Blank Error – No action taken
100681 Juneau PD 1: (1) Ambient Fail – No action taken
100683 AST/Glennallen: (1) Blank Error, (2) User Aborts – No action taken
100685 UAA PD: (1) Internal Standard Error – No action taken
100687 Bethel PD: (3) Tank Interference, (1) Blank Error – Instrument removed from service 1/27/2015
100693 Valdez PD: (1) Blank Error – No action taken