

CALIBRATION FACTORS

Alaska Department of Public Safety
DATAMASTER dmt: 100367

Date: 09/25/2019
Time: 05:37:55

OPERATOR NAME:
Q Q Q
CERTIFICATE #: Q

Ca = 0.1000
CAL = 0.998594 0.800 <= CAL < 1.200
b1 = 0.0003 0.0000 <= b1 < 0.0040
b2 = 0.0042 0.0010 <= b2 < 0.0100
b3 = 0.0004 0.0000 <= b3 < 0.0040
Xq = 0.0824 0.0500 <= Xq < 0.2000
a21 = 1.172415 1.050 <= a21 < 1.300
a31 = 0.403874 0.300 <= a31 < 0.800