

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100665

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 04/10/2015

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 25.61: 0.068
LOT #: 11313080A1
EXPIRATION: 06/01/2015
TANK PRESSURE: 1054 psi

BLANK TEST	0.000	14:52
INTERNAL STANDARD	VERIFIED	14:52
EXTERNAL STANDARD	0.068	14:52
BLANK TEST	0.000	14:53
EXTERNAL STANDARD	0.067	14:54
BLANK TEST	0.000	14:54
EXTERNAL STANDARD	0.068	14:55
BLANK TEST	0.000	14:56
EXTERNAL STANDARD	0.068	14:56
BLANK TEST	0.000	14:57
EXTERNAL STANDARD	0.068	14:57
BLANK TEST	0.000	14:58

Average = 0.0678
Std Dev = 0.0004

Diagnostic Check

VERSIONS

DMT: 3.00
PIC: 3.03
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.9°C PASSED
Breath Tube = 48.1°C PASSED

PUMP INFO

Flow Rate = 4.494 L/M PASSED

DETECTOR INFO

PUMP ON PASSED
PUMP OFF PASSED

FILTER INFO


Filter 1 PASSED
Filter 2 PASSED
Filter 3 PASSED

INTERNAL STANDARD

PASSED

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.
- (5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.


Nita J. Bolz
Scientific Director
State Breath Alcohol Program

5/15/2015

Subscribed and sworn before me this 15th day of May, 20 15


Nikki Roth, Notary Public
My Commission Expires With Office



BMB
4/15/15