

# VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100346

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 03/27/2019

## External Standard Test Values

### EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080  
TARGET AT 30.12: 0.081  
LOT #: AG826101  
EXPIRATION: 09/18/2020  
TANK PRESSURE: 668 psi

BLANK TEST	0.000	21:29
INTERNAL STANDARD	VERIFIED	21:29
EXTERNAL STANDARD	0.080	21:29
BLANK TEST	0.000	21:30
EXTERNAL STANDARD	0.080	21:30
BLANK TEST	0.000	21:31
EXTERNAL STANDARD	0.080	21:31
BLANK TEST	0.000	21:32
EXTERNAL STANDARD	0.080	21:33
BLANK TEST	0.000	21:33
EXTERNAL STANDARD	0.080	21:34
BLANK TEST	0.000	21:35

Average = 0.0800  
Std Dev = 0.0000

## Diagnostic Check

### VERSIONS

DMT: 3.02  
PIC: 3.03  
Modem: 2.6  
Questions: 2.2

### TEMPERATURES

Sample Chamber = 48.7°C PASSED  
Breath Tube = 47.8°C PASSED

### PUMP INFO

Flow Rate = 4.535 L/M PASSED

### DETECTOR INFO

PUMP ON PASSED  
PUMP OFF PASSED


### FILTER INFO

Filter 1 PASSED  
Filter 2 PASSED  
Filter 3 PASSED

INTERNAL STANDARD PASSED


I, Charles R. Foster, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.
- (5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

  
Charles R. Foster  
Scientific Director  
State Breath Alcohol Program

5/31/19

Subscribed and sworn before me this 31 day of MAY, 20 19

  
Notary Public  
My Commission Expires With Office



JW 05/07/19