

Alaska Scientific Crime Detection Laboratory

DataMaster Check In Form

Issued: 1/13/2017
Effective: 1/17/2017

Version: DMCIF 2017 R0
Status: Active

Instrument S/N 100397

Received From USBPD

Date 6/28/21

Analyst C. O'Bryant

1. Reason for instrument return:

Preventative Maintenance

2. Change instrument and case location in JusticeTrax Breath Alcohol Database

3. Ensure change in instrument status form is in JusticeTrax

4. Upload Instrument would not upload.

5. Update agency information and polling status in DM Host Calling Program

6. Check for VOCs

From Intoximeters

~~7. Intoximeters Repair Paperwork (attached)~~ ^{NA}

Changed the modem board, but still wasn't able to get the instrument to upload. COB 7/13/21