

Alaska Scientific Crime Detection Laboratory  
Change in Instrument Status Form

Version: 4.0

Effective: 8/10/2022

Date: 8/15/2023

Instrument #: 100421

Instrument Location: 8517 20<sup>th</sup> St Joint Base Elemendorf Richardson

Supervisor Name and Agency: Scott Gardner/6222/JBR4

Describe the Status Change or Issue with the Instrument:

DATAMASTER 100421 removed from service after several ambient fails and a standard out of range status message. Delivered to crime lab per request.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: [Click or tap here to enter text.](#)

Out of Service Date/Time: 0800/9 Aug 2023

-----  
**For Use by SCDL**

Additional Notes

[Click or tap here to enter text.](#)

**Email completed form to [dps.scdl.toxicology@alaska.gov](mailto:dps.scdl.toxicology@alaska.gov)**  
For questions contact the Breath Alcohol Program at 907-269-5740

Page 1 of 1

All printed copies are uncontrolled.

Approved by: Chemistry Supervisor