

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100682

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 11/06/2019

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 29.99: 0.080
LOT #: AG735001
EXPIRATION: 12/16/2019
TANK PRESSURE: 385 psi

BLANK TEST	0.000	17:55
INTERNAL STANDARD	VERIFIED	17:55
EXTERNAL STANDARD	0.078	17:55
BLANK TEST	0.000	17:56
EXTERNAL STANDARD	0.078	17:56
BLANK TEST	0.000	17:57
EXTERNAL STANDARD	0.078	17:57
BLANK TEST	0.000	17:58
EXTERNAL STANDARD	0.078	17:58
BLANK TEST	0.000	17:59
EXTERNAL STANDARD	0.079	17:59
BLANK TEST	0.000	18:00

Average = 0.0782
Std Dev = 0.0004

Diagnostic Check

VERSIONS
DMT: 3.02
PIC: 3.03
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.9°C PASSED
Breath Tube = 38.5°C PASSED

PUMP INFO

Flow Rate = 4.716 L/M PASSED

DETECTOR INFO

PUMP ON PASSED
PUMP OFF PASSED


FILTER INFO

Filter 1 PASSED
Filter 2 PASSED
Filter 3 PASSED

INTERNAL STANDARD PASSED


I, Brandi M. Barnett, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.
- (5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.


 12/13/19

Brandi M. Barnett
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 13th day of Dec. 20 19



Notary Public
My Commission Expires With Office



AM 12/9/19