

Alaska Scientific Crime Detection Laboratory

Change in Instrument Status Form

Issued: 6/11/2014
Effective: 6/11/2014

Version: CIISF 2014 R0
Status: Active

Date: 6/9/17 Instrument Serial Number: 100694

Instrument Location: Petersburg Police Department, 14 S. Nordic Dr., Petersburg Alaska 99833

Supervisor Name and Agency: Derek Thorsen, Petersburg Police Department

Describe the Status Change or Issue with the Instrument:

Instrument was taken out of service because of an incomplete VOC. Interference was detected due to building construction. Instrument was taken out of service on 6/6/17 at 1948hrs and returned to service on 6/9/17 at 0138 hrs. after valid VOC was completed. No further action is warranted.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: 6/9/17 at 0138 hrs

Out of Service Date/Time: 6/6/17 at 1948 hrs.

For Use by SCDL

Additional Notes

Email completed form to dps.scdl.toxicology@alaska.gov

For questions contact the Breath Alcohol Program at 907-269-5740