

# Alaska Scientific Crime Detection Laboratory

## Change in Instrument Status Form

Issued: 6/11/2014  
Effective: 6/11/2014

Version: CIISF 2014 R0  
Status: Active

Date: 7/5/17 Instrument Serial Number: 100413

Instrument Location: Cooper Landing / AST

Supervisor Name and Agency: Colleen O'Bryant, SCOD

Describe the Status Change or Issue with the Instrument:

\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: \_\_\_\_\_

Out of Service Date/Time: 7/5/17

For Use by SCDL

Additional Notes

Removed from service to test standard out of range low (status message environmental versus analytical). Removed by BTO Trooper Tim Lewis at my direction

COB 7/5/17

Email completed form to [dps.scdl.toxicology@alaska.gov](mailto:dps.scdl.toxicology@alaska.gov)

For questions contact the Breath Alcohol Program at 907-269-5740

## **O'Bryant, Colleen S (DPS)**

---

**From:** O'Bryant, Colleen S (DPS)  
**Sent:** Wednesday, July 05, 2017 12:01 PM  
**To:** Foster, Charles R (DPS); brandi.barnett@alaska.gov  
**Subject:** AST/Cooper Landing

**Cooper Landing is having issues with external standards out of range low. I gave Trooper Lewis a new tank to try, it was still read out of range low. Brandi stopped by the post to troubleshoot the instrument and was unable to pinpoint if the status messages were environmental or not. Trooper Lewis replaced the instrument on 07/05/2017 (with my assistance and instructions as the Cooper Landing Post does not currently have a DataMaster Supervisor) and still received standard out of range messages on an attempted VOC. Trooper Lewis indicated the Cooper Landing post was small, a garage and office area attached. There is not currently anyone stationed at that post and the airflow and traffic in and out of the post has decreased.**

**I spoke with Sgt Zweifel and he will attempt to increase the airflow at the post over the next week and continue troubleshooting. Trooper Lewis is going to return 100413 (original instrument) to the lab.**

**Colleen O'Bryant – Forensic Scientist III  
Blood and Breath Alcohol  
Alaska Scientific Crime Detection Laboratory  
4805 Dr. Martin Luther King Jr. Ave  
Anchorage, AK 99507**

**907-269-5783 (direct)  
907-350-2111 (cell)**