

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100698

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 01/19/2017

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 29.47: 0.079
LOT #: AG608504
EXPIRATION: 03/25/2018
TANK PRESSURE: 1207 psi

BLANK TEST	0.000	21:42
INTERNAL STANDARD	VERIFIED	21:42
EXTERNAL STANDARD	0.081	21:42
BLANK TEST	0.000	21:43
EXTERNAL STANDARD	0.080	21:43
BLANK TEST	0.000	21:44
EXTERNAL STANDARD	0.080	21:44
BLANK TEST	0.000	21:45
EXTERNAL STANDARD	0.080	21:45
BLANK TEST	0.000	21:46
EXTERNAL STANDARD	0.081	21:46
BLANK TEST	0.000	21:47

Average = 0.0804
Std Dev = 0.0005

Diagnostic Check

VERSIONS

DMT: 3.02
PIC: 3.03
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.9°C PASSED
Breath Tube = 46.2°C PASSED

PUMP INFO

Flow Rate = 4.717 L/M PASSED

DETECTOR INFO

PUMP ON PASSED
PUMP OFF PASSED

FILTER INFO

Filter 1 PASSED
Filter 2 PASSED
Filter 3 PASSED

INTERNAL STANDARD PASSED


I, Charles R. Foster, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.
- (5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

 3/10/17

Charles R. Foster
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 10th day of Mar, 2017


Nikki Roth, Notary Public
My Commission Expires With Office



00132/04/17