

	State of Alaska Scientific Crime Detection Laboratory		
	DataMaster Calibration Form - DMCAF2011R1		
	Issuing Authority - Nita Bolz, Scientific Director		
	Approval date: 03/10/11	Effective date: 03/10/11	Archive date: Active

DataMaster DMT Calibration Form

Instrument # 100889Date: 4/12/11

1. Wet Bath Simulator Solution

- a. Level 0.100
- b. Manufacturer Cuth
- c. Lot 10260
- d. Expiration Date 10/26/12

2. Wet Bath Simulator

- a. Serial Number DR4161
- b. Date Calibrated 8/23/10

3. Attach Calibration Factors

Breath Alcohol Analyst C. Bryant

CALIBRATION FACTORS

Alaska Department of Public Safety
DATAMASTER dmt: 100389

Date: 04/12/2011
Time: 14:51:47

SUPERVISOR NAME:
COLLEEN S OBRYANT
CERTIFICATE #: 6237

LOT #: GUYH100
BOTTLE #: 10260
EXPIRATION: 10/25/2012

Ca	=	0.1000			
CAL	=	0.989059	0.800	<=	CAL < 1.200
b1	=	0.0001	0.0000	<=	b1 < 0.0040
b2	=	0.0045	0.0010	<=	b2 < 0.0100
b3	=	0.0002	0.0000	<=	b3 < 0.0040
Xq	=	0.1009	0.0500	<=	Xq < 0.2000
a21	=	1.173159	1.050	<=	a21 < 1.300
a31	=	0.422527	0.300	<=	a31 < 0.800

COB 2/2

Final page for report

\

\\ANCSCLSAN01\JTRAX\CASEDOCS\B-100389\B-100389_ECF-20140

2/19/2014