

Alaska Scientific Crime Detection Laboratory

DataMaster Acceptance Form

Issued: 01/27/2012
Effective: 01/27/2012

Version: DMAF2012 R0
Status: Active

Instrument # 100681

Date: 5/11/12

Analyst: CRF

- Check in instrument in to JusticeTrax Breath Alcohol Database
- Visual Inspection
- Update software *USB Already Fixed*
- Print calibration factors (attached)
- NPAS calibration paperwork (attached)
- Barcode Instrument
- State of Alaska Property ID# 12-F0372

Already vented to rec, controller board new



CALIBRATION FACTORS

Alaska Department of Public Safety
DATAMASTER dmt: 100681

Date: 10/25/2011
Time: 11:27:48

Ca = 0.1000
CAL = 0.999423 0.800 <= CAL < 1.200
b1 = 0.0000 0.0000 <= b1 < 0.0040
b2 = 0.0023 0.0010 <= b2 < 0.0100
b3 = 0.0000 0.0000 <= b3 < 0.0040
Xq = 0.1073 0.0500 <= Xq < 0.2000
a21 = 1.171221 1.050 <= a21 < 1.300
a31 = 0.478868 0.300 <= a31 < 0.800



Certificate of Calibration

100681
Instrument Serial Number

10260
Standard Reference Number

The above instrument has been calibrated and is hereby certified for use in determining alcohol breath concentration when used in accordance with the operating guidelines by a properly trained operator.

10-26-2011
Date

[Signature]
Technician

Final page for report

\

\\ANCSCLSAN01\JTRAX\CASEDOCS\B-100681\B-100681 ECF-20140

7/10/2014