

Alaska Scientific Crime Detection Laboratory

Change in Instrument Status Form

Issued: 6/11/2014
Effective: 6/11/2014

Version: CIISF 2014 R0
Status: Active

Date: 4/7/15 Instrument Serial Number: 100351

Instrument Location: ARD Jail

Supervisor Name and Agency: Colleen O'Bryant / SCDL

Describe the Status Change or Issue with the Instrument:

I was at the Anchorage Jail troubleshooting another instrument when I noticed the breath hose had been removed from 100351.
I found two breath hoses stored in a cabinet drawer close by.
I was unable to determine which one was the original breath hose so I removed the instrument from service.
The last few tests performed resulted in gas flow errors, Tank psi 247.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: ~~4/6/15~~ 4/6/15 08:46 08:41/15

Out of Service Date/Time: 4/6/15 ~ 12pm

For Use by SCDL

Additional Notes

Email completed form to dps.scdl.toxicology@alaska.gov

For questions contact the Breath Alcohol Program at 907-269-5740

SUBJECT TEST

Alaska Department of Public Safety
DATA MASTER dmt: 100351

Date: 04/06/2015
Time: 04:48:24

SUBJECT NAME:
TEST TEST
LICENSE #: 0

OPERATOR NAME:
G WITTE
OPERATOR'S #: 4949

DEPT/AGENCY:
ANC1
CASE/REPORT: 0
TEST TYPE: TEST

EXTERNAL STANDARD INFORMATION
NOMINAL: 0.080
TARGET AT 29.53: 0.079
LOT #: 09214080A2
EXPIRATION: 05/01/2016

DIAGNOSTIC CHECK PASSED 04:49
BLANK TEST 0.000 04:49
INTERNAL STANDARD VERIFIED 04:49

Gas Flow Error