

Alaska DMT Software Revision 3.00 Evaluation

December 3, 2013

100397 was used in the evaluation

1. Loaded 3.00 software and verified it can be pushed through the modem lines.
 - a. Took ~4.5 minutes to push through the modem lines
 - b. When poll disconnected the DMT restarted status was "VOC required"
 - c. Started VOC test with Operator password
2. VOC ran successfully
 - a. #5 from request list and email 5/21/2013
tank pressure in psi printed under expiration date
 - b. #6 from request list
barometric pressure calculation of target value changed to round to the third digit instead of truncating
 $\text{psi } 30.53/29.92 = 0.0816$
VOC listed target at 0.082 correctly
 - c. Verified software was 3.00
3. #12 from request list – verified in software "about" the next scheduled VOC date was listed and was correct at 12/12/2013 12:00:00
 - a. Ran a second VOC and ensure the next scheduled VOC was still correct at 12/12/13
 - b. Changed the date of the DMT to 12/12/2013 manually to allow the regularly scheduled VOC to start. DMT started the VOC. The scheduled VOC was updated in the software to the 1/4/2014 date correctly.
4. From email dated 3/22/2013 verified that "supervisor test" is no longer an option in the software with a supervisor password and set up menu options have also been removed from the supervisor password. Verified that lab personnel have access to supervisor test option.
5. Ran a non-drinking complete subject test.
 - a. #7 from request list the volume printed on the subject sample line
 - b. #9 from the request list the minimum flow line was printed at 3.00 L/min
6. #2 4th bullet from request list. Ran a non-drinking subject test with 10 attempts. The test stopped after the 10th attempt was made.
 - a. Confirmed the wording of the status message "Maximum Attempts Exceeded" and the wording "The test has been aborted due to too many attempts. Instruct the subject to provide a sample with one long, steady breath."
7. # 3 from the request list - Any time a status message occurs generate a pop up window immediately the status message pop up box that says "A status message has occurred and the test is not complete. Would you like to test the subject

again?" Then have a yes or no box. If yes is selected print the test strip and return to the data entry screen with the previous data entered. If no is selected print the test strip and return to the home screen.

- a. Confirmed with maximum number of attempts – yes and returned to data entry screen
 - b. Confirmed with incomplete – no and returned to home screen
 - c. Confirmed with RFI triggered during the subject sample, confirmed status message box, no, confirmed signature box and driver's license card reader.
 - d. Confirmed with RFI triggered after subject sample, confirmed status message box, no, confirmed signature box.
 - e. Confirmed with Invalid subject sample, confirmed status message box, confirmed "XXX" was subject test result on screen.
8. # 4 from request list if the subject test does not complete after the subject has blown the test strip will print with a subject test result of "XXX"
- a. Confirmed with an RFI triggered after the subject sample was accepted.
 - b. Confirmed with a blank error after the subject sample was accepted.
9. # 2 1st bullet from request list. Ran Interference detected on the subject sample. Confirmed the status message box and confirmed "XXX" result.
10. #2 3rd bullet from request list. Triggered RFI during the subject sample. Confirmed status message box.
11. #2 2nd bullet from request list.
- a. Triggered a blank error and confirmed the status message box.
 - b. Triggered an Interference Detected during the subject sample and confirmed status message.

NDJ/aly
2/5/2015

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100397

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 12/03/2013

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080

TARGET AT 30.53: 0.082 2b

LOT #: 22612080A1

EXPIRATION: 09/01/2014

TANK PRESSURE: 678 psi 2a

BLANK TEST	0.000	09:11
INTERNAL STANDARD	VERIFIED	09:11
EXTERNAL STANDARD	0.080	09:12
BLANK TEST	0.000	09:12
EXTERNAL STANDARD	0.080	09:13
BLANK TEST	0.000	09:14
EXTERNAL STANDARD	0.080	09:14
BLANK TEST	0.000	09:15
EXTERNAL STANDARD	0.080	09:15
BLANK TEST	0.000	09:16
EXTERNAL STANDARD	0.080	09:16
BLANK TEST	0.000	09:17

Average = 0.0800

Std Dev = 0.0000

Diagnostic Check

VERSIONS

DMT: 3.00 2c

PIC: 3.02

Modem: 2.6

Questions: 2.2

TEMPERATURES

Sample Chamber = 48.8°C PASSED

Breath Tube = 44.2°C PASSED

PUMP INFO

Flow Rate = 4.238 L/M PASSED

DETECTOR INFO

PUMP ON PASSED

PUMP OFF PASSED

FILTER INFO

Filter 1 PASSED

Filter 2 PASSED

Filter 3 PASSED

INTERNAL STANDARD PASSED

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

(1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.

(2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.

(3) I am the Scientific Director of the State Breath Alcohol Program.

(4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.

(5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program.

(6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz

Scientific Director

State Breath Alcohol Program

Subscribed and sworn before me this _____ day of _____, 20 _____

SUBJECT TEST

Alaska Department of Public Safety
 DATA MASTER dmt: 100397

Date: 12/03/2013
 Time: 09:57:32

SUBJECT NAME:
 TEST TEST
 LICENSE #: NONDRINKING


OPERATOR NAME:
 BRANDI BARNETT
 OPERATOR'S #: 7038

DEPT/AGENCY:
 AAA1
 CASE/REPORT: XXXX
 TEST TYPE: TEST

EXTERNAL STANDARD INFORMATION
 NOMINAL: 0.080
 TARGET AT 30.54: 0.082
 LOT #: 22612080A1
 EXPIRATION: 09/01/2014

DIAGNOSTIC CHECK	PASSED	09:58
BLANK TEST	0.000	09:58
INTERNAL STANDARD	VERIFIED	09:58
EXTERNAL STANDARD	0.080	09:58
BLANK TEST	0.000	09:59
SUBJECT SAMPLE (Vol=1.56L)	0.000	10:00
BLANK TEST	0.000	10:01
EXTERNAL STANDARD	0.080	10:01
BLANK TEST	0.000	10:02

5a



SUBJECT TEST

Alaska Department of Public Safety
 DATAMASTER dmt: 100397

Date: 12/03/2013
 Time: 10:11:25

SUBJECT NAME:
 TEST TEST
 LICENSE #: MIN # ATMPPTS

leal
 7/a

OPERATOR NAME:
 COLLEEN S O'BRYAN
 OPERATOR'S #: 6237

DEPT/AGENCY:
 AAA1
 CASE/REPORT: 13-1234
 TEST TYPE: TEST


EXTERNAL STANDARD INFORMATION
 NOMINAL: 0.080
 TARGET AT 30.54: 0.082
 LOT #: 22612080A1
 EXPIRATION: 09/01/2014

DIAGNOSTIC CHECK	PASSED	10:12
BLANK TEST	0.000	10:12
INTERNAL STANDARD	VERIFIED	10:12
EXTERNAL STANDARD	0.080	10:12
BLANK TEST	0.000	10:13
SUBJECT SAMPLE	XXX	10:14

Maximum Attempts Exceeded

- Confirmed Max Attempt Message
- Confirmed not completed test message
- Confirmed signature bar present
- Confirmed press [OK] to test subject again takes you to data entry screen
- Confirmed status message triggered after 10 attempts

CCB



SUBJECT TEST

Alaska Department of Public Safety
 DATA MASTER dmt: 100397

Date: 12/03/2013
 Time: 10:17:00

SUBJECT NAME:
 TEST TEST
 LICENSE #: INCOMPLETE

OPERATOR NAME:
 COLLEEN S OBRYANT
 OPERATOR'S #: 6237


DEPT/AGENCY:
 AAA1
 CASE/REPORT: 13-1234
 TEST TYPE: TEST

EXTERNAL STANDARD INFORMATION
 NOMINAL: 0.080
 TARGET AT 30.54: 0.082
 LOT #: 22612080A1
 EXPIRATION: 09/01/2014

DIAGNOSTIC CHECK	PASSED	10:17
BLANK TEST	0.000	10:18
INTERNAL STANDARD	VERIFIED	10:18
EXTERNAL STANDARD	0.080	10:18
BLANK TEST	0.000	10:19
SUBJECT SAMPLE	INCOMPLETE	10:21

*Confirmed status message box
 (no change to wording)
 confirmed press [ue] to not
 test the subject again.*

BSMB



SUBJECT TEST

Alaska Department of Public Safety
 DATAMASTER dmt: 100397

Date: 12/03/2013
 Time: 10:44:42

SUBJECT NAME:
 TEST TEST
 LICENSE #: AK7298726

OPERATOR NAME:
 BRANDI BARNETT
 OPERATOR'S #: 7038

DEPT/AGENCY:
 RFI BEFORE SUBJECT SAMPLE
 CASE/REPORT: XXXX
 TEST TYPE: TEST

EXTERNAL STANDARD INFORMATION
 NOMINAL: 0.080
 TARGET AT 30.54: 0.082
 LOT #: 22612080A1
 EXPIRATION: 09/01/2014

DIAGNOSTIC CHECK
 BLANK TEST
 INTERNAL STANDARD
 EXTERNAL STANDARD
 BLANK TEST


RFI DETECTED

PASSED 10:46
 0.000 10:46
 VERIFIED 10:46
 0.081 10:46
 0.000 10:47

- RFI during subject test*
- *Confirmed status message box*
 - *and RFI status message appearing*
 - *confirmed signature box*
 - *Confirmed drivers license card reader*

BMB

7c \$10



SUBJECT TEST

Alaska Department of Public Safety
 DATA MASTER dmt: 100397

Date: 12/03/2013
 Time: 10:52:15

SUBJECT NAME:
 TEST TEST
 LICENSE #: AK7298726

OPERATOR NAME:
 BRANDI BARNETT
 OPERATOR'S #: 7038

DEPT/AGENCY:
 RFI AFTER SUBJECT SAMPLE
 CASE/REPORT: XXXX
 TEST TYPE: TEST

EXTERNAL STANDARD INFORMATION
 NOMINAL: 0.080
 TARGET AT 30.54: 0.082
 LOT #: 22612080A1
 EXPIRATION: 09/01/2014


DIAGNOSTIC CHECK	PASSED	10:53
BLANK TEST	0.000	10:53
INTERNAL STANDARD	0.080	10:53
EXTERNAL STANDARD	0.080	10:53
BLANK TEST	0.000	10:54
SUBJECT SAMPLE (Vol=2.31L)	XXX	10:55
BLANK TEST	0.000	10:56

RFI DETECTED

- Confirmed xxx to subject test on test with status msg following accepted sample.
- Confirmed status message bar and RFI status message bar.
- Confirmed signature bar

[Handwritten Signature]

Alcohol(g/210L)
 Flow Rate(L/M)
 Min. Flow Rate(L/M)



SUBJECT TEST

Alaska Department of Public Safety
DATA MASTER dmt: 100397

Date: 12/03/2013
Time: 11:00:53

SUBJECT NAME:
TEST TEST
LICENSE #: AK7298726

OPERATOR NAME:
BRANDI BARNETT
OPERATOR'S #: 7038

DEPT/AGENCY:
INVALID
CASE/REPORT: XXXX
TEST TYPE: TEST


EXTERNAL STANDARD INFORMATION
NOMINAL: 0.080
TARGET AT 30.54: 0.082
LOT #: 22612080A1
EXPIRATION: 09/01/2014

DIAGNOSTIC CHECK		PASSED	
BLANK TEST	0.000	11:01	11:02
INTERNAL STANDARD	0.080	11:02	11:02
EXTERNAL STANDARD	0.000	11:02	11:03
BLANK TEST	0.000	11:03	11:03
SUBJECT SAMPLE	INVALID	11:03	11:03

- o confirmed invalid status message
- the + retest adding
- o confirmed the for subject test result on the screen.
- o confirmed signature the

(Handwritten signature)

Alcohol(g/210L)
Flow Rate(L/M)
Min. Flow Rate(L/M)



SUBJECT TEST

Alaska Department of Public Safety
 DATA MASTER dmt: 100397

Date: 12/03/2013
 Time: 13:26:35

SUBJECT NAME:
 TEST TEST
 LICENSE #: XXXX

OPERATOR NAME:
 BRANDI BARNETT
 OPERATOR'S #: 7038

DEPT/AGENCY:
 BLANK ERROR AFTER TEST
 CASE/REPORT: XXXX
 TEST TYPE: TEST


EXTERNAL STANDARD INFORMATION
 NOMINAL: 0.080
 TARGET AT 30.52: 0.082
 LOT #: 22612080A1
 EXPIRATION: 09/01/2014

DIAGNOSTIC CHECK	PASSED	13:27
BLANK TEST	0.000	13:27
INTERNAL STANDARD	VERIFIED	13:27
EXTERNAL STANDARD	0.080	13:28
BLANK TEST	0.000	13:29
SUBJECT SAMPLE (Vol=2.17L)	XXX	13:29

Blank Error

- Confirmed xx for subject test result with non-completed test
- Confirmed blank error following subject test

BMV



SUBJECT TEST

Alaska Department of Public Safety
DATAMASTER dmt: 100397

Date: 12/03/2013
Time: 13:20:46

SUBJECT NAME:
TEST TEST
LICENSE #: INTERFERENCE

OPERATOR NAME:
BRANDI BARNETT
OPERATOR'S #: 7038

DEPT/AGENCY:
AAAI
CASE/REPORT: XXXX
TEST TYPE: TEST


EXTERNAL STANDARD INFORMATION
NOMINAL: 0.080
TARGET AT 30.52: 0.082
LOT #: 22612080A1
EXPIRATION: 09/01/2014

DIAGNOSTIC CHECK	
BLANK TEST	PASSED 13:22
INTERNAL STANDARD	0.000 13:22
EXTERNAL STANDARD	VERIFIED 13:22
BLANK TEST	0.080 13:22
SUBJECT SAMPLE	0.000 13:23
	XXX 13:24

Interference Detected

- Confirmed Interference message
- by
- confirmed rx for non-completed
- subject tests,
- confirmed signature block

SLT



SUBJECT TEST

Alaska Department of Public Safety
DATAMASTER dmt: 100397

Date: 12/03/2013
Time: 11:05:26

SUBJECT NAME:
TEST TEST
LICENSE #: AK7298726

OPERATOR NAME:
BRANDI BARNETT
OPERATOR'S #: 7038


DEPT/AGENCY:
STANDARD OUT OF RANGE AFTER
CASE/REPORT: XXXX
TEST TYPE: TEST

EXTERNAL STANDARD INFORMATION
NOMINAL: 0.080
TARGET AT 30.54: 0.082
LOT #: 22612080A1
EXPIRATION: 09/01/2014

DIAGNOSTIC CHECK	PASSED	11:06
BLANK TEST	0.000	11:06
INTERNAL STANDARD	VERIFIED	11:06
EXTERNAL STANDARD	0.080	11:06
BLANK TEST	0.000	11:07
SUBJECT SAMPLE (Vol=2.07L)	XXX	11:08
BLANK TEST	0.000	11:09
EXTERNAL STANDARD	0.100	11:09

Standard Out of Range

Handwritten signature/initials.



SUBJECT TEST

Alaska Department of Public Safety
 DATA MASTER dmt: 100397

Date: 12/03/2013
 Time: 13:46:12

SUBJECT NAME:
 TEST TEST
 LICENSE #: XXXX

OPERATOR NAME:
 BRANDI BARNETT
 OPERATOR'S #: 7038


DEPT/AGENCY:
 REFUSAL QUERY
 CASE/REPORT: XXXX
 TEST TYPE: TEST

EXTERNAL STANDARD INFORMATION
 NOMINAL: 0.080
 TARGET AT 30.52: 0.082
 LOT #: 22612080A1
 EXPIRATION: 09/01/2014

DIAGNOSTIC CHECK	PASSED	13:47
BLANK TEST	0.000	13:47
INTERNAL STANDARD	VERIFIED	13:47
EXTERNAL STANDARD	0.080	13:47
BLANK TEST	0.000	13:48
SUBJECT SAMPLE	INCOMPLETE	13:50

*Refusal Query on
 No to refusal
 Confirmed result*

BMB



SUBJECT TEST

Alaska Department of Public Safety
DATAMASTER dmt: 100397

Date: 12/03/2013
Time: 13:40:22

SUBJECT NAME:
TEST TEST
LICENSE #: XXXX

OPERATOR NAME:
BRANDI BARNETT
OPERATOR'S #: 7038

DEPT/AGENCY:
REFUSAL QUERY
CASE/REPORT: XXXX
TEST TYPE: TEST

EXTERNAL STANDARD INFORMATION
NOMINAL: 0.080
TARGET AT 30.52: 0.082
LOT #: 22612080A1
EXPIRATION: 09/01/2014

DIAGNOSTIC CHECK	PASSED	13:41
BLANK TEST	0.000	13:41
INTERNAL STANDARD	VERIFIED	13:41
EXTERNAL STANDARD	0.080	13:41
BLANK TEST	0.000	13:42
SUBJECT SAMPLE	REFUSED	13:42
BLANK TEST	0.000	13:43

*Refusal every on
yes to refusal
continued result*

BBB

Alaska Software Revision (3.00)

Scott below are the details that we have recorded included in the 3.00 Software change that we began with you in early 2013. We are now ready to finalize the software revision and are preparing to roll it into use in January of 2015. Can you please verify some parts of the software changes (areas in blue below) in writing for us since they are not things we can verify ourselves.

1. Change ambient fail parameters:

- a. Reduce allowable change at 2 and 20 seconds from 20 mV to 12 mV **done**
- b. Possibly change 20 second reading to later in purge **not changed**

Were these two changes made to the ambient fail parameters?

2. When these status message occurs (at any time during the test sequence) , add a statement to the existing pop up box for the following:

- Interference Detected on subject sample (not on the external stds)
"Test subject a second time, if a second test results in an INTERFERENCE message, obtain a legal blood sample" **done**
- Blank error and ambient fail
"Move the subject and any source of air contamination away from the DMT. Ensure rear of the DMT is not blocked." **done**
- RFI
"Ensure all radios in the vicinity of the DMT are off" **done**
- EXCEEDED ATTEMPTS
After 10 attempts in the two minute window have a pop up with the wording " The test has been aborted due to too many attempts. Instruct the subject to provide a sample with one long, steady breath. "
done

3. Any time a status message occurs generate a pop up window immediately the status message pop up box that says "A status message has occurred and the test is not complete. Would you like to test the subject again?" Then have a yes or no box. If yes is selected print the test strip and return to the data entry screen with the previous data entered. If no is selected print the test strip and return to the home screen. **done**

4. If the subject test does not complete after the subject has blown print the test strip with the subject test result as XXX. **done**

5. Add tank pressure to VOC – print under expiration date of cylinder **done**

6. barometric pressure calculation of target value change to round the third digit instead of truncate (all the time) **done**

Is this change in place?

7. Print the volume on the subject sample **done**

8. Compare final result with last data point? **done**

Was this change made and please detail the explanation? The last point on the graph and the static analysis reading are compared to see if they are within 0.005 of each other. If they do not agree then a Static Analysis Error occurs. (Status #40)

9. Graph with minimum flow line / change cutoff to 3.0 L/Min instead 2.78

Is the minimum rate 3.0 or 2.78? 3.0

10. Fix standard deviation to 0.0030

Has this change been made? done

11. Fix user abort- modem issues **Please remind me what this was**

12. Add next VOC date into about **done**

Additional information that we also need:

Status code for "Exceeded Attempts" **#101**

Status code for "Datapoint comparison" **is this Static Analysis Error? If so, then #40**

Definition of an "attempt" **3 readings above 3L/M flow rate**

Does the specific breath hose used affect the RFI sensitivity of that DMT? **yes**

Was there a printer driver update? **The driver that permits pdf creation is the current one.**

Bolz, Nita J (DPS)

From: Barnett, Brandi M (DPS)
Sent: Friday, March 22, 2013 9:44 AM
To: eng@npas.com
Cc: Bolz, Nita J (DPS)
Subject: Software 3.00

Scott,

#4 ✓ ✓

Would it be possible to remove the supervisor test menu options for the breath test supervisors in our software. We would like to gray out the setup option for them and not allow them to run one. We just taught a class and they are very confused by it because they don't use it.

Also, I was never able to get the remote DMT program to work on our computer so could we troubleshoot this next week. Monday is a holiday for the state but I will be in on Tuesday.

Brandi Barnett
Forensic Scientist III
Alaska Scientific Crime Detection Laboratory
(907)269-5695 Phone

Bolz, Nita J (DPS)

From: Barnett, Brandi M (DPS)
Sent: Tuesday, February 05, 2013 12:33 PM
To: eng@npas.com
Cc: Bolz, Nita J (DPS)
Subject: Alaska Software Revision 3.00 Change

Scott,

I spoke with Nita and she decided that we will leave the averaging of linearity and supervisor test as it currently is where it averages from the three decimals. Let me know if you have any other questions.

Brandi Barnett
Forensic Scientist III
Alaska Scientific Crime Detection Laboratory
(907)269-5695 Phone

Bolz, Nita J (DPS)

From: Scott Marhefka <eng@npas.com>
Sent: Thursday, February 05, 2015 12:56 PM
To: Bolz, Nita J (DPS)
Subject: Re: Alaska 3.00 software

Yes, this is in the new software.

Scott

On 2/5/2015 4:47 PM, Bolz, Nita J (DPS) wrote:

Hi Scott,

I am writing up my final paperwork on this software change and had one final question for you.

On the "fix user abort-modem issue" this was about when we timed out using the modem lines sometimes it would cause a user abort. Just need confirmation this is in the new software.

Thanks,

Nita

--

Scott Marhefka
National Patent Analytical Systems
2090 Harrington Memorial Road
Mansfield, OH 44903

email: eng@npas.com
phone: (419) 526-6727