

Alaska Scientific Crime Detection Laboratory

Change in Instrument Status Form

Issued: 6/11/2014
Effective: 6/11/2014

Version: CIISF 2014 R0
Status: Active

Date: 9/8/16 Instrument Serial Number: 100405

Instrument Location: Clear AFB

Supervisor Name and Agency: Colleen O'Bryant

Describe the Status Change or Issue with the Instrument:

Instrument removed from service remotely to
confirm lot number on external standard.
Lot number was confirmed (09214080123) and
tank was reinstalled with a VOC performed.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: 8/3/2014

Out of Service Date/Time: _____

For Use by SCDL

Additional Notes

Email completed form to dps.scdl.toxicology@alaska.gov

For questions contact the Breath Alcohol Program at 907-269-5740