

Alaska Scientific Crime Detection Laboratory

Change in Instrument Status Form

Issued: 12/15/2017
Effective: 12/15/2017

Version: CIISF 2017 R0
Status: Active

Date: 03/04/20 Instrument Serial Number: 100694

Instrument Location: APD Jail

Supervisor Name and Agency: Dereen Walton (SCDL) on the phone with Patrick Martin (APD)

Describe the Status Change or Issue with the Instrument:

Instrument was not uploading. Requested Ofc. Martin to check phone line. Ofc. Martin found that the modem line had been ripped out of the wall and was damaged. Ofc. Martin borrowed a functional line from another DMT at the jail to allow 100694 to upload. Ofc. Martin indicated that he would locate additional modem cords to replace the damaged one and have spares. 100694 uploaded successfully.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: NA

Out of Service Date/Time: NA

For Use by SCDL

Additional Notes

Email completed form to dps.scdl.toxicology@alaska.gov

For questions contact the Breath Alcohol Program at 907-269-5740