

Alaska Scientific Crime Detection Laboratory

Change in Instrument Status Form

Issued: 6/11/2014
Effective: 6/11/2014

Version: CIISF 2014 R0
Status: Active

Date: 6/7/16 Instrument Serial Number: 100375

Instrument Location: Seldovia PD

Supervisor Name and Agency: Brandi Barnett

Describe the Status Change or Issue with the Instrument:

Instrument was unable to perform tests after 5/1/16 due to external standard expiration. No operators at location and Troopers are not interested in maintaining. Instrument will be removed from service by Eugene Fowler.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: _____

Out of Service Date/Time: 5/1/16

For Use by SCDL

Additional Notes

Email completed form to dps.scdl.toxicology@alaska.gov

For questions contact the Breath Alcohol Program at 907-269-5740