

Alaska Scientific Crime Detection Laboratory
Change in Instrument Status Form

Version: 4.0

Effective: 8/10/2022

Date: 7/21/2023

Instrument #: 100355

Instrument Location: Homer Police Department

Supervisor Name and Agency: Trooper David Bozman/AST

Describe the Status Change or Issue with the Instrument:

Instrument removed from service due to Filter Wheel Error and on-site trouble shooting was unsuccessful.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: N/A

Out of Service Date/Time: 07/21/2023 1500

For Use by SCDL

Additional Notes

I had Trooper Bozman open the DataMaster and jiggle/press down on the spring located on top of the solenoid tube to ensure that it's not stuck and causing the filter wheel error. This did not fix the issue. So a replacement instrument was sent. TLM 07/24/2023

Email completed form to dps.scdl.toxicology@alaska.gov
For questions contact the Breath Alcohol Program at 907-269-5740

Page 1 of 1

All printed copies are uncontrolled.

Approved by: Chemistry Supervisor