

# VERIFICATION OF CALIBRATION REPORT

*of DataMaster dmt Breath Test Instrument*

*State of Alaska*

Serial #: 100678

*Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*

Date: 01/02/2023

## External Standard Test Values

### EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080  
TARGET AT 28.79: 0.077  
LOT #: AG135004  
EXPIRATION: 12/16/2023  
TANK PRESSURE: 1136 psi

BLANK TEST	0.000	10:43
INTERNAL STANDARD	VERIFIED	10:44
EXTERNAL STANDARD	0.076	10:44

Blank Error

## Diagnostic Check

### VERSIONS

DMT: 3.02  
PIC: 3.03  
Modem: 2.6  
Questions: 2.2

### TEMPERATURES

Sample Chamber = 48.9°C	PASSED
Breath Tube = 45.4°C	PASSED

### PUMP INFO

Flow Rate = 4.825 L/M	PASSED
-----------------------	--------

### DETECTOR INFO

PUMP ON	PASSED
PUMP OFF	PASSED

### FILTER INFO

Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

INTERNAL STANDARD	PASSED
-------------------	--------

File Only DMV 02/10/23