

Alaska Scientific Crime Detection Laboratory

DataMaster Check In Form

Issued: 1/13/2017
Effective: 1/17/2017

Version: DMCIF 2017 R0
Status: Active

Instrument S/N 100691

Received From Denali National Park

Date 05/03/21

Analyst Derek Watton

1. Reason for instrument return:

Screen frozen, instrument not placed in service at Denali NP

2. Change instrument and case location in JusticeTrax Breath Alcohol Database

3. Ensure change in instrument status form is in JusticeTrax

4. Upload

*DJW 05/03/21
Instrument never placed in service. No tests
run at Denali NP. DJW 05/03/21*

5. Update agency information and polling status in DM Host Calling Program

6. Check for VOCs

From Intoximeters

7. Intoximeters Repair Paperwork (attached)

Instrument was previously sent to the manufacturer for freezing in September, 2020. Instrument will be returned to manufacturer for warranty repair. DJW 05/06/2021