

# VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100352

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 09/04/2018

## External Standard Test Values

### EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080  
TARGET AT 30.12: 0.081  
LOT #: AG634001  
EXPIRATION: 09/04/2018\*  
TANK PRESSURE: 1010 psi

BLANK TEST	0.000	21:06
INTERNAL STANDARD	VERIFIED	21:06
EXTERNAL STANDARD	0.080	21:06
BLANK TEST	0.000	21:07
EXTERNAL STANDARD	0.080	21:08
BLANK TEST	0.000	21:08
EXTERNAL STANDARD	0.080	21:09
BLANK TEST	0.000	21:10
EXTERNAL STANDARD	0.080	21:10
BLANK TEST	0.000	21:11
EXTERNAL STANDARD	0.080	21:11
BLANK TEST	0.000	21:12

Average = 0.0800  
Std Dev = 0.0000

## Diagnostic Check

### VERSIONS

DMT: 3.02  
PIC: 3.03  
Modem: 2.6  
Questions: 2.2

### TEMPERATURES

Sample Chamber = 48.8°C PASSED  
Breath Tube = 41.5°C PASSED

### PUMP INFO

Flow Rate = 5.086 L/M PASSED

### DETECTOR INFO

PUMP ON PASSED  
PUMP OFF PASSED

### FILTER INFO

Filter 1 PASSED  
Filter 2 PASSED  
Filter 3 PASSED

INTERNAL STANDARD PASSED

Wrong expiration date  
File only COB 10/2/18  
✓