

Alaska Scientific Crime Detection Laboratory

DataMaster Check In Form

Effective: 10/10/2022

Version: 3.0

Instrument #: 100699

Received From: JBEP

Date: 10/7/24

Analyst: Kaija Swisher

Reason for instrument return:

Due to number of internal standard errors on subject tests from IFIR of 7/19/24-9/5/24, it was determined to remove the instrument from service.

Change instrument and case location in JusticeTrax Breath Alcohol Database

Ensure change in instrument status form is in JusticeTrax

Upload

Update agency information and polling status in DMHost Calling Program

Check for VOCs

From Intoximeters

Intoximeters Repair Paperwork (attached)