

Alaska Scientific Crime Detection Laboratory

DataMaster Check In Form

Issued: 7/10/2015
Effective: 7/10/2015

Version: DMCIF 2015 R1
Status: Active

Instrument S/N 100344

Received From King Cove PD

Date 7/9/17

Analyst Colleen Bryant

1. Reason for instrument return:

Didn't upload in the field, brought back to evaluate.

2. Change instrument and case location in JusticeTrax Breath Alcohol Database

3. Ensure change in instrument status form is in JusticeTrax

4. Upload (Permetek box was loose)

5. Update agency information and polling status in DM Host Calling Program NA

6. Check for VOCs

(Instrument returning to agency)

From NPAS

7. NPAS Repair Paperwork (attached)