

Alaska Scientific Crime Detection Laboratory

DataMaster Check In Form

Issued: 2/2/2015
Effective: 2/2/2015

Version: DMCIF 2015 R0
Status: Active

Instrument S/N 100666

Received From ARD

Date 3/20/15

Analyst Colleen O'Bryant

1. Reason for instrument return:

Removed from service by Nita Botz for difficulty restarting following C&D software update.

2. Change instrument and case location in JusticeTrax Breath Alcohol Database

3. Ensure change in instrument status form is in JusticeTrax

4. Visual Inspection

5. Upload

6. Update agency information and polling status in DM Host Calling Program

7. Check for VOCs

From NPAS

8. NPAS Repair Paperwork (attached)

VERIFICATION OF CALIBRATION REPORT*of DataMaster dmt Breath Test Instrument**State of Alaska*

Serial #: 100666

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 03/20/2015

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
 TARGET AT 29.35: 0.078
 LOT #: 09214080A2
 EXPIRATION: 05/01/2016
 TANK PRESSURE: 1000 psi

BLANK TEST	0.000	14:20
INTERNAL STANDARD	VERIFIED	14:20
EXTERNAL STANDARD	0.076	14:20
BLANK TEST	0.000	14:21
EXTERNAL STANDARD	0.076	14:21
BLANK TEST	0.000	14:22
EXTERNAL STANDARD	0.076	14:22
BLANK TEST	0.000	14:23
EXTERNAL STANDARD	0.076	14:23
BLANK TEST	0.000	14:24
EXTERNAL STANDARD	0.076	14:24
BLANK TEST	0.000	14:25

Average = 0.0760
 Std Dev = 0.0000

Diagnostic Check

VERSIONS

DMT: 3.00
 PIC: 3.03
 Modem: 2.6
 Questions: 2.2

TEMPERATURES

Sample Chamber = 49.0°C PASSED
 Breath Tube = 47.2°C PASSED

PUMP INFO

Flow Rate = 4.934 L/M PASSED

DETECTOR INFO

PUMP ON PASSED
 PUMP OFF PASSED

FILTER INFO

Filter 1 PASSED
 Filter 2 PASSED
 Filter 3 PASSED

INTERNAL STANDARD PASSED

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.
- (5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

 Nita J. Bolz
 Scientific Director
 State Breath Alcohol Program

Subscribed and sworn before me this ____ day of _____, 20 ____