

# Alaska Scientific Crime Detection Laboratory

## Change in Instrument Status Form

Issued: 12/15/2017  
Effective: 12/15/2017

Version: CIISF 2017 R0  
Status: Active

Date: 6/17/21 Instrument Serial Number: 100374

Instrument Location: Delta Det.

Supervisor Name and Agency: AST Sergeant Jason Pug

Describe the Status Change or Issue with the Instrument:

---

---

---

---

---

---

---

---

---

---

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: 6/17/21 / 1350 hrs

Out of Service Date/Time: \_\_\_\_\_

For Use by SCDL

Additional Notes

---

---

---

---

---

---

---

---

---

---

Email completed form to [dps.scdl.toxicology@alaska.gov](mailto:dps.scdl.toxicology@alaska.gov)

For questions contact the Breath Alcohol Program at 907-269-5740