

# VERIFICATION OF CALIBRATION REPORT

*of DataMaster dmt Breath Test Instrument*

*State of Alaska*

Serial #: 100687

*Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*

Date: 04/08/2025

## External Standard Test Values

## Diagnostic Check

### VERSIONS

DMT: 3.02

PIC: 3.02

Modem: 2.6

Questions: 2.2

### TEMPERATURES

Sample Chamber = 48.9°C

PASSED

Breath Tube = 45.5°C

PASSED

### PUMP INFO

Flow Rate = 4.376 L/M

PASSED

### DETECTOR INFO

PUMP ON

FAILED

PUMP OFF

PASSED

### FILTER INFO

Filter 1

PASSED

Filter 2

PASSED

Filter 3

PASSED

### INTERNAL STANDARD

PASSED

Diagnostic Check Failed

File Only  
CW 5/12/25