

# VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100420

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 06/11/2015

## External Standard Test Values

### EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080  
TARGET AT 28.53: 0.076  
LOT #: 09214080A3  
EXPIRATION: 05/01/2016  
TANK PRESSURE: 769 psi

BLANK TEST	0.000	10:39
INTERNAL STANDARD	VERIFIED	10:39
EXTERNAL STANDARD	0.076	10:39
BLANK TEST	0.000	10:40
EXTERNAL STANDARD	0.074	10:41
BLANK TEST	0.000	10:41
EXTERNAL STANDARD	0.075	10:42
BLANK TEST	0.000	10:43
EXTERNAL STANDARD	0.075	10:43
BLANK TEST	0.000	10:44
EXTERNAL STANDARD	0.075	10:44
BLANK TEST	0.000	10:45

Average = 0.0750  
Std Dev = 0.0007

## Diagnostic Check

### VERSIONS

DMT: 3.01  
PIC: 3.02  
Modem: 2.6  
Questions: 2.2

### TEMPERATURES

Sample Chamber = 48.7°C PASSED  
Breath Tube = 45.7°C PASSED

### PUMP INFO

Flow Rate = 4.765 L/M PASSED

### DETECTOR INFO

PUMP ON PASSED  
PUMP OFF PASSED


### FILTER INFO

Filter 1 PASSED  
Filter 2 PASSED  
Filter 3 PASSED


INTERNAL STANDARD PASSED


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.
- (5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 12<sup>th</sup> day of June 20 15

  
Nikki Roth, Notary Public  
My Commission Expires With Office



003 6/11/15