

VERIFICATION OF CALIBRATION REPORT

*of DataMaster dmt Breath Test Instrument
State of Alaska*

Serial #: 100374

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 10/10/2019

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 29.51: 0.079
LOT #: AG826101
EXPIRATION: 09/18/2020
TANK PRESSURE: 158 psi

BLANK TEST	0.000	20:34
INTERNAL STANDARD	VERIFIED	20:34
EXTERNAL STANDARD	0.077	20:34
BLANK TEST	0.000	20:35
EXTERNAL STANDARD	INTERFERENCE	20:35

Interference Detected

Diagnostic Check

VERSIONS

DMT: 3.02
PIC: 3.02
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber =	48.6°C	PASSED
Breath Tube =	45.9°C	PASSED

PUMP INFO

Flow Rate =	4.246 L/M	PASSED
-------------	-----------	--------

DETECTOR INFO

PUMP ON	PASSED
PUMP OFF	PASSED

FILTER INFO

Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

INTERNAL STANDARD

PASSED