

Alaska Scientific Crime Detection Laboratory

DataMaster Check In Form

Effective: 10/10/2022

Version: 3.0

Instrument #: 100682

Received From: Intox

Date: 10/23/23

Analyst: Charlie Glaister

Reason for instrument return:

Returned from Intox. Was sent due to filter wheel B interference status messages.

Change instrument and case location in JusticeTrax Breath Alcohol Database

~~Ensure change in instrument status form is in JusticeTrax~~ See 100682 2023.06.19 Repair for Return to Factory Info.

~~Upload~~ CWB 10/23/23
No agency info to upload. Modem did connect. CWB 10/23/23

Update agency information and polling status in DMHost Calling Program

~~Check for VOCs~~ CWB 10/23/23
No agency VOCs to check.

From Intoximeters

Intoximeters Repair Paperwork (attached)

SRO Number: SRO-057718

Received Date: 07/05/2023

Completion Date: 08/30/2023

SRO Type: REPAIR

SRO Description: DMT Repair

Customer ID: C000003651

Customer PO: 07052023-OBryant

Contact: Colleen OBryant

Email: Colleen.obryant@alaska.gov

Ship To:

Alaska Scientific Crime Detection Laboratory
4805 Dr MLK Jr Ave
Anchorage AK 99507
USA

Units on SRO

1	LEGDMT	LEGACY DMT
---	--------	------------

Service Line: 1		**All Instruments are Calibrated to Factory Specifications**	
Unit Type:	(LEGDMT) LEGACY DMT		
Serial Number:	50100682		
Reason:	DMT Status Messages		
	FltWl er		
Reason Notes:	Customer Reported Issue: Filter wheel issue		
Resolution:	Replaced		
	RPL comp		
Resolution Notes:	Rebuilt filter wheel assembly. Replaced filter wheel solenoid. Replaced chopper motor. Performed filter wheel test, 59:0 pass		
	1	59-0000-07	DMT HRDW WASHER SLEEVE RETAINER 39611
Reason:	DMT Status Messages		
	Blank		
Reason Notes:	Detector found to be unstable during standard service. Detectors, when unstable, are contributors to Blank Errors.		
Resolution:	Replaced		
	RPL comp		
Resolution Notes:	Replaced detector. Adjusted instrument to factory specifications.		
	1	59-0080-00	DMT ELEC ASSY SOLENOID FILTR WHEEL 41344
	1	59-0006-00	DMT HRDW ASSY MOTOR 12X20.6MM 41310
	1	59-0008-00	DMT HRDW COLLAR-CHOPPER WHEEL 31433
	1	59-0055-00	DMT ELEC DETECTOR PCB ASSY 41601
	1	59-0083-00	DMT ELEC LAMP ASSY IR TO -8 41630

Order Packing Slip Report

Intoximeters
Experience • Service • Integrity

10/9/2023 1:46:08 PM

Packing Slip: 291957

Pack Date: 10/9/2023

SRO: SRO-057718



From: Intoximeters, Inc
2081 Craig Road
St Louis MO 63146
USA

Customer: C000003651

Ship To: 1

Alaska Department of Public Safety
4805 Dr MLK Jr Ave
Anchorage AK 99507
USA

Alaska Scientific Crime Detection
Laboratory
4805 Dr MLK Jr Ave
Anchorage AK 99507
USA

Customer PO: 07052023-OBryant

Ship Via: FDX

FED EX

Weight: 37.00

Delivery Terms: 2D

Shipping Terms: ORIGIN

Packages: 1

Customer Ship Account:

Line Number	Item	UM	Ordered Quantity	Shipment Quantity	Backorder Quantity
1	LEGDMT	LEGACY DMT			50100682
1	59-0000-07	EA		1.00	0.00
	DMT HRDW WASHER SLEEVE RETAINER 39611				
2	59-0080-00	EA		1.00	0.00
	DMT ELEC ASSY SOLENOID FILTR WHEEL 41344				
3	59-0006-00	EA		1.00	0.00
	DMT HRDW ASSY MOTOR 12X20.6MM 41310				
4	59-0008-00	EA		1.00	0.00
	DMT HRDW COLLAR-CHOPPER WHEEL 31433				
5	59-0055-00	EA		1.00	0.00
	DMT ELEC DETECTOR PCB ASSY 41601				
6	59-0083-00	EA		1.00	0.00
	DMT ELEC LAMP ASSY IR TO -8 41630				