

Alaska Scientific Crime Detection Laboratory

DataMaster Acceptance Form

Issued: 01/27/2012
Effective: 01/27/2012

Version: DMAF2012 R0
Status: Active

Instrument # 100673

Date: 4/9/12

Analyst: CRF

- Check in instrument in to JusticeTrax Breath Alcohol Database
- Visual Inspection
- Update software
- Print calibration factors (attached)
- NPAS calibration paperwork (attached)
- Barcode Instrument
- State of Alaska Property ID# 12-F0364

New controller board present and exhaust is plumbed to back



CALIBRATION FACTORS

Alaska Department of Public Safety
DATAMASTER dmt: 100673

Date: 10/27/2011
Time: 02:04:26

Ca	=	0.1000		
CAL	=	0.991994	0.800	<= CAL < 1.200
b1	=	0.0004	0.0000	<= b1 < 0.0040
b2	=	0.0046	0.0010	<= b2 < 0.0100
b3	=	0.0001	0.0000	<= b3 < 0.0040
Xq	=	0.0755	0.0500	<= Xq < 0.2000
a21	=	1.215480	1.050	<= a21 < 1.300
a31	=	0.457053	0.300	<= a31 < 0.800



Certificate of Calibration

100673

Instrument Serial Number

10260

Standard Reference Number

The above instrument has been calibrated and is hereby certified for use in determining alcohol breath concentration when used in accordance with the operating guidelines by a properly trained operator.

10-27-11

Date



Technician