

Alaska Scientific Crime Detection Laboratory

DataMaster Check In Form

Effective: 3/20/2025

Version: 4.0

Instrument #: 100687

Received From: King Cove PD

Date: 11/17/25

Analyst: Kaya Swisher

Reason for instrument return:

Internal standard errors.

From the field

Change instrument and case location in JusticeTrax Breath Alcohol Database ☒

Ensure change in instrument status form is in JusticeTrax ☒

Upload ☒

Update agency information and polling status in DMHost Calling Program ☒

Check for VOCs ☒

From Intoximeters

Intoximeters Repair Paperwork (attached) ☐

Change instrument and case location in JusticeTrax Breath Alcohol Database ☐

Reset Options and Clear Memory ☐