

# VERIFICATION OF CALIBRATION REPORT

*of DataMaster dmt Breath Test Instrument*

*State of Alaska*

Serial #: 100687

*Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*

Date: 04/12/2019

## External Standard Test Values

### EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080  
TARGET AT 29.36: 0.079  
LOT #: AG735001  
EXPIRATION: 12/16/2019  
TANK PRESSURE: 613 psi

BLANK TEST	0.000	03:41
INTERNAL STANDARD	VERIFIED	03:41
EXTERNAL STANDARD	0.075	03:42
BLANK TEST	0.000	03:43
EXTERNAL STANDARD	0.076	03:43
BLANK TEST	0.000	03:44
EXTERNAL STANDARD	0.076	03:44
BLANK TEST	0.000	03:45
EXTERNAL STANDARD	0.073	03:45

Blank Error

## Diagnostic Check

### VERSIONS

DMT: 3.02  
PIC: 3.02  
Modem: 2.6  
Questions: 2.2

### TEMPERATURES

Sample Chamber = 48.7°C	PASSED
Breath Tube = 41.2°C	PASSED

### PUMP INFO

Flow Rate = 4.376 L/M	PASSED
-----------------------	--------

### DETECTOR INFO

PUMP ON	PASSED
PUMP OFF	PASSED

### FILTER INFO

Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

INTERNAL STANDARD	PASSED
-------------------	--------