

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100376

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 10/24/2021

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080

TARGET AT 28.94: 0.077

LOT #: ~~AG-0102~~ **AG-018102** **CKG-1118/21** *

EXPIRATION: 06/29/2022

TANK PRESSURE: 1192 psi

BLANK TEST	0.000	19:54
INTERNAL STANDARD	VERIFIED	19:54
EXTERNAL STANDARD	0.077	19:54
BLANK TEST	0.000	19:55
EXTERNAL STANDARD	0.076	19:55
BLANK TEST	0.000	19:56
EXTERNAL STANDARD	0.076	19:56
BLANK TEST	0.000	19:57
EXTERNAL STANDARD	0.077	19:58
BLANK TEST	0.000	19:58
EXTERNAL STANDARD	0.077	19:59
BLANK TEST	0.000	20:00

Average = 0.0766

Std Dev = 0.0005

* Incorrect lot #.

See 100376 2021.11.08 Change In Status
CKG-1119/21

Diagnostic Check

VERSIONS

DMT: 3.02

PIC: 3.02

Modem: 2.6

Questions: 2.2

TEMPERATURES

Sample Chamber = 48.8°C PASSED

Breath Tube = 46.3°C PASSED

PUMP INFO

Flow Rate = 3.854 L/M PASSED

DETECTOR INFO

PUMP ON PASSED

PUMP OFF PASSED

FILTER INFO

Filter 1 PASSED

Filter 2 PASSED

Filter 3 PASSED

INTERNAL STANDARD PASSED

I, Brandi M. Barnett, after being first duly sworn, depose and state as follows:

(1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.

(2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.

(3) I am the Scientific Director of the State Breath Alcohol Program.

(4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.


(5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program.

(6) The referenced instrument is certified for evidentiary use in the State of Alaska.


 12/3/21

Brandi M. Barnett
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 03 day of 12, 2021



Notary Public
My Commission Expires With Office



CKG-1118/21