

# Alaska Scientific Crime Detection Laboratory

## Change in Instrument Status Form

Issued: 12/15/2017  
Effective: 12/15/2017

Version: CIISF 2017 R0  
Status: Active

Date: 11-12-2019 Instrument Serial Number: 100413

Instrument Location: HOMER P.D.

Supervisor Name and Agency: Jim Knott HOMER P.D.

Describe the Status Change or Issue with the Instrument:

Subject Sample w/ no resistance followed by interference and  
blank error tests. Checked hose attachment checked also  
3 additional tests show no sir flow.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: \_\_\_\_\_

Out of Service Date/Time: 11-12-2019 0848

For Use by SCDL

Additional Notes

Five way valve sticking. JM 11/12/19

Email completed form to [dps.scdl.toxicology@alaska.gov](mailto:dps.scdl.toxicology@alaska.gov)

For questions contact the Breath Alcohol Program at 907-269-5740