	State of Alaska Scientific Crime Detection Laboratory		
	DataMaster Acceptance Form – DMAF2011R0		
	Issuing Authority – Nita Boiz, Scientific Director		
Approval date: 03/01/11	Effective date: 03/01/11	Archive date: Active	

DataMaster DMT Acceptance Form

Instrument # 100670

Date: 10/14/11

Analyst: CRF

1. Check in instrument in to JusticeTrax Breath Alcohol Database
2. Visual Inspection
3. Update software
4. Print calibration factors (attached)
5. NPAS calibration paperwork (attached)
6. Barcode Instrument
7. State of Alaska Property ID# 12-F0362

**B-100670**      100670



QSHS03C40W0MVV      B-100670


Breath Program

CALIBRATION FACTORS

-----  
Alaska Department of Public Safety  
DATAMASTER dmt: 100670  
-----

Date: 09/21/2011  
Time: 14:17:27

Ca	=	0.1500		
CAL	=	0.983698	0.800	<= CAL < 1.200
b1	=	0.0002	0.0000	<= b1 < 0.0040
b2	=	0.0023	0.0010	<= b2 < 0.0100
b3	=	0.0000	0.0000	<= b3 < 0.0040
Xq	=	0.0886	0.0500	<= Xq < 0.2000
a21	=	1.179698	1.050	<= a21 < 1.300
a31	=	0.439665	0.300	<= a31 < 0.800



**National Patent  
Analytical Systems, Inc.**

*Certificate of Calibration*

100670

Instrument Serial Number

11030

Standard Reference Number

The above instrument has been calibrated and is hereby certified for use in determining alcohol breath concentration when used in accordance with the operating guidelines by a properly trained operator.

9/22/11

Date

Charles E. Townsend

Technician

Final page for report

\

\\ANCSCLSAN01\JTRAX\CASEDOCS\B-100670\B-100670 ECF-20140

7/10/2014