

Alaska Scientific Crime Detection Laboratory

DataMaster Check In Form

Effective: 11/10/2021

Version: 2.0

Instrument #: 100395

Received From: Fairbanks Airport PD

Date: 25 Aug 2022

Analyst: Tatiana Malnoske

Supervising Analyst: Carlie Glaister

Reason for instrument return:

Elevated internal standard values.

Change instrument and case location in JusticeTrax Breath Alcohol Database

Ensure change in instrument status form is in JusticeTrax

Upload

Update agency information and polling status in DMHost Calling Program

Check for VOCs

From Intoximeters

Intoximeters Repair Paperwork (attached)