

Alaska Scientific Crime Detection Laboratory

DataMaster Evaluation Form

Issued: 7/10/2015
Effective: 7/10/2015

Version: DMEF2015 R0
Status: Active

Instrument # 100687

Date: 8/24/15

Analyst: Colleen O'Bryen

1. Reason for instrument evaluation

Attempted certification and instrument displayed "undefined error"

2. Visual Inspection N/A

3. Troubleshooting/Repair

Attempted to roll software back to 2.00 and reinstall in the following order: 2.00, 3.00, 3.01. (Still getting "undefined error" at initiation) of VOC.

4. VOC (attached) N/A

5. Internal Standard value $\leq 3.00\%$ Yes No N/A

6. Verify software version 3.01

Return to Factory for Repair

Ready For Calibration

Ready For Certification