

Alaska Scientific Crime Detection Laboratory  
Change in Instrument Status Form

Version: 4.0

Effective: 8/10/2022

Date: 12/27/2022

Instrument #: 100345

Instrument Location: North Pole PD

Supervisor Name and Agency: Carlie Glaister

Describe the Status Change or Issue with the Instrument:

During IFIR I noticed a higher than average prevalence of RFI status messages. I emailed Breath Test Supervisor Jed Smith notifying him and asking him to remind officers to not key radios within the vicinity of the DataMaster.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: NA

Out of Service Date/Time: NA

-----  
**For Use by SCDL**

Additional Notes

Click or tap here to enter text.

**Email completed form to [dps.scdl.toxicology@alaska.gov](mailto:dps.scdl.toxicology@alaska.gov)**  
For questions contact the Breath Alcohol Program at 907-269-5740

Page 1 of 1

All printed copies are uncontrolled.

Approved by: Chemistry Supervisor

## Glaister, Carlie K (DPS)

---

**From:** Jed Smith <JSmith@northpolepolice.org>  
**Sent:** Tuesday, December 27, 2022 8:42 AM  
**To:** Glaister, Carlie K (DPS)  
**Subject:** Re: RFI

**CAUTION:** This email originated from outside the State of Alaska mail system. Do not click links or open attachments unless you recognize the sender and know the content is safe.

Copy, thanks!

V/R,



Lieutenant Jed Smith  
North Pole Police Dept.  
125 Snowman Lane  
North Pole, Alaska 99705  
Office: 907-488-6902  
Desk: 907-488-8459

---

**From:** Glaister, Carlie K (DPS) <carlie.glaister@alaska.gov>  
**Sent:** Tuesday, December 27, 2022 8:40:16 AM  
**To:** Jed Smith  
**Subject:** RFI

Hello, I am reviewing status messages for the last VOC cycle, and North Pole PD has a higher than average prevalence of RFI status messages. Please remind officers to not key radios within the vicinity of the DataMaster.

*Carlie Glaister (Bailey)*

Forensic Scientist III – Chemistry Section  
State of Alaska Scientific Crime Detection Laboratory  
907-269-5689