

# VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100345

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 04/28/2020

## External Standard Test Values

### EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080  
TARGET AT 29.34: 0.078  
LOT #: AG826101  
EXPIRATION: 09/18/2020  
TANK PRESSURE: 560 psi

|                   |          |       |
|-------------------|----------|-------|
| BLANK TEST        | 0.000    | 14:38 |
| INTERNAL STANDARD | VERIFIED | 14:38 |
| EXTERNAL STANDARD | 0.075    | 14:38 |
| BLANK TEST        | 0.000    | 14:39 |
| EXTERNAL STANDARD | 0.075    | 14:40 |
| BLANK TEST        | 0.000    | 14:40 |
| EXTERNAL STANDARD | 0.075    | 14:41 |
| BLANK TEST        | 0.000    | 14:42 |
| EXTERNAL STANDARD | 0.075    | 14:42 |
| BLANK TEST        | 0.000    | 14:43 |
| EXTERNAL STANDARD | 0.075    | 14:43 |
| BLANK TEST        | 0.000    | 14:44 |

Average = 0.0750  
Std Dev = 0.0000

## Diagnostic Check

VERSIONS  
DMT: 3.02  
PIC: 3.02  
Modem: 2.6  
Questions: 2.2

### TEMPERATURES

Sample Chamber = 48.8°C PASSED  
Breath Tube = 46.5°C PASSED

### PUMP INFO

Flow Rate = 4.146 L/M PASSED

### DETECTOR INFO

PUMP ON PASSED  
PUMP OFF PASSED


### FILTER INFO

Filter 1 PASSED  
Filter 2 PASSED  
Filter 3 PASSED

INTERNAL STANDARD PASSED


I, Brandi M. Barnett, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.
- (5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

 5/12/20

Brandi M. Barnett  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 12 day of May 20 20

  
Notary Public  
My Commission Expires With Office



CAB 4/28/20