

CALIBRATION FACTORS

Alaska Department of Public Safety
DATAMASTER dmt: 100417

Date: 03/25/2020
Time: 05:59:25

OPERATOR NAME:
INTOX I METERS
CERTIFICATE #: 1

Ca = 0.2000	
CAL = 0.972760	0.800 <= CAL < 1.200
b1 = 0.0000	0.0000 <= b1 < 0.0040
b2 = 0.0044	0.0010 <= b2 < 0.0100
b3 = 0.0000	0.0000 <= b3 < 0.0040
Xq = 0.0941	0.0500 <= Xq < 0.2000
a21 = 1.147054	1.050 <= a21 < 1.300
a31 = 0.516811	0.300 <= a31 < 0.800