

# Alaska Scientific Crime Detection Laboratory

## DataMaster Acceptance Form

Issued: 01/27/2012  
Effective: 01/27/2012

Version: DMAF2012 R0  
Status: Active

Instrument # 100682

Date: 9/18/12


Analyst: Brandi Barnett

- Check in instrument in to JusticeTrax Breath Alcohol Database
- Visual Inspection
- Update software  *N/A broken when opened won't get to home screen*
- Print calibration factors (attached)  *N/A*
- NPAS calibration paperwork (attached)
- Barcode Instrument
- State of Alaska Property ID# 12-FO 373

**B-100682**



*Return to NPAs 9/18/12*



**National Patent  
Analytical Systems, Inc.**

*Certificate of Calibration*

100687

Instrument Serial Number


10260

Standard Reference Number

The above instrument has been calibrated and is hereby certified for use in determining alcohol breath concentration when used in accordance with the operating guidelines by a properly trained operator.

10-25-2011

Date

  
Technician

Final page for report

\

\\ANCSCLSAN01\JTRAX\CASEDOCS\B-100682\B-100682 ECF-20140

7/10/2014