

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100395

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 01/13/2020

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 29.96: 0.080
LOT #: AG826101
EXPIRATION: 09/18/2020
TANK PRESSURE: 1214 psi

BLANK TEST	0.000	09:41
INTERNAL STANDARD	VERIFIED	09:42
EXTERNAL STANDARD	0.080	09:42
BLANK TEST	0.000	09:43
EXTERNAL STANDARD	0.079	09:43
BLANK TEST	0.000	09:44
EXTERNAL STANDARD	0.079	09:44
BLANK TEST	0.000	09:45
EXTERNAL STANDARD	0.079	09:45
BLANK TEST	0.000	09:46
EXTERNAL STANDARD	0.079	09:46
BLANK TEST	0.000	09:47

Average = 0.0792
Std Dev = 0.0004

Diagnostic Check

VERSIONS

DMT: 3.02
PIC: 3.02
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.9°C PASSED
Breath Tube = 47.3°C PASSED

PUMP INFO

Flow Rate = 4.274 L/M PASSED

DETECTOR INFO

PUMP ON PASSED
PUMP OFF PASSED


FILTER INFO

Filter 1 PASSED
Filter 2 PASSED
Filter 3 PASSED

INTERNAL STANDARD PASSED


I, Brandi M. Barnett, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.
- (5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.



 1/30/20

Brandi M. Barnett
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 30th day of Jan, 2020



Notary Public
My Commission Expires With Office



003/01/2020