

Infield Instrument Review 1/4/2019 to 2/21/2019

Prepared by Brandi Barnett

100346: Anchorage PD- 2/57 Tank Interference- No action taken

100347: Craig PD- 4/11 Gas Flow Error- Instrument removed from service

100348: Ft. Wainwright- 1/4 Pump Error- No action taken

100351: Fairbanks Airport- 1/4 RFI- No action taken

100354: Bristol Bay Borough- 1/3 Pump Error- No action taken

100355: Seward PD- 3/6 RFI- Monitor

100361: Bethel PD- 1/7 RFI- No action taken

100374: Anchorage PD- 1/122 Ambient Fail- No action taken

100378: AST/Fairbanks- 1/42 RFI- No action taken

100379: Wasilla PD- 3/27 RFI, 1/27 Pump Error- No action taken

100382: UAA- 1/10 Pump Error, 1/10 Tank Interference- No action taken

100386: Unalakleet PD- 1/3 Pump Error- No action taken

100389: Cordova PD- 2/13 RFI- No action taken

100405: Dillingham PD- 1/8 Tank Interference, 2/8 Blank Error- Remove from Service

100415: Sitka PD- 1/11 Tank Interference, 1/11 Blank Error, 1/11 Internal Standard Error- Remove from service

100416: Fairbanks PD- 7/60 RFI, 4/60 Tank Interference- No action taken

100418: Kenai PD- 1/16 RFI, 1/16 Pump Error, 1/16 Blank Error, 1/16 Diagnostic Failed- Diagnostic failed and Blank error same subject- No action taken

100668: Palmer PD- 1/28 Pump Error, 2/28 Blank Error- No action taken

100678: AST Mat-Su West- 5/31 RFI, 1/31 Interference, 1/31 Tank Interference- No action taken

100687: AST Soldotna- 1/10 Tank Interference, 1/10 Blank Error- No action taken

100690: Unalaska- 3/8 Standard Out of Range- Tank was changed on 1/19 no further problems, No action taken.

100694: Anchorage PD- 2/99 Pump Error, 1/99 Blank Error- No action taken

100695: JBER- 2/13 Pump Error, 1/13 Interference- No action taken

Does not include complete data for Ft. Wainwright PD.

Recommend Tank replacement for AST Seldovia and AST Angoon due to tank reading low on VOC.



