

# Alaska Scientific Crime Detection Laboratory

## Change in Instrument Status Form

Issued: 12/15/2017  
Effective: 12/15/2017

Version: CIISF 2017 R0  
Status: Active

Date: 9-30-18 Instrument Serial Number: 100682

Instrument Location: Heppner Bay

Supervisor Name and Agency: Malvin Colley, AST

Describe the Status Change or Issue with the Instrument:

taken out of service and being packed + sent  
back to crime lab per their request

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: \_\_\_\_\_

Out of Service Date/Time: 9-30-18 @ 1400

For Use by SCDL

Additional Notes

Internal standard errors, detector overflow on VOCs 9/10/18

Email completed form to [dps.scdl.toxicology@alaska.gov](mailto:dps.scdl.toxicology@alaska.gov)

For questions contact the Breath Alcohol Program at 907-269-5740