

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100369

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 07/23/2024

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 29.96: 0.080
LOT #: AG310901
EXPIRATION: 04/19/2025
TANK PRESSURE: 976 psi

BLANK TEST	0.000	10:08
INTERNAL STANDARD	VERIFIED	10:09
EXTERNAL STANDARD	0.076	10:09
BLANK TEST	0.000	10:10
EXTERNAL STANDARD	0.078	10:10
BLANK TEST	0.000	10:11
EXTERNAL STANDARD	0.078	10:11
BLANK TEST	0.000	10:12
EXTERNAL STANDARD	INTERFERENCE	10:12

Interference Detected

Diagnostic Check

VERSIONS

DMT: 3.02
PIC: 3.03
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.9°C	PASSED
Breath Tube = 48.1°C	PASSED

PUMP INFO

Flow Rate = 4.682 L/M	PASSED
-----------------------	--------

DETECTOR INFO

PUMP ON	PASSED
PUMP OFF	PASSED

FILTER INFO

Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

INTERNAL STANDARD

PASSED

File Only DM 08/21/24