

# Alaska Scientific Crime Detection Laboratory

## Change in Instrument Status Form

Issued: 12/15/2017  
Effective: 12/15/2017

Version: CIISF 2017 R0  
Status: Active

Date: Nov 6th, 2019 Instrument Serial Number: 100682

Instrument Location: Sitka, AK

Supervisor Name and Agency: Gary W. Cranford / Sitka PD

### Describe the Status Change or Issue with the Instrument:

On 11/6/19, I received an email from Crime Lab regarding poor conductivity with DMT.  
Initially was unable to activate touch screen and turned off power to DMT, which resolved the issue.  
After powering up DMT, completed Diagnostic Record check, which resulted in all functions PASSED.  
1st attempted VOC resulted in a Blank Error.  
2nd attempted at VOC resulted in Blank Error.  
Completed a purge of the sample chamber for approximately 5 minutes due to the two Blank Errors.  
3rd attempt at VOC was successful.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: N/A

Out of Service Date/Time: N/A

### For Use by SCDL

### Additional Notes

Email from the crime lab was due to ~~a~~ connectivity through  
the modem line. Connectivity was reestablished. BMB 11/27/19

Email completed form to [dps.scdl.toxicology@alaska.gov](mailto:dps.scdl.toxicology@alaska.gov)

For questions contact the Breath Alcohol Program at 907-269-5740