

Alaska Scientific Crime Detection Laboratory

DataMaster Check In Form

Issued: 1/13/2017
Effective: 1/17/2017

Version: DMCIF 2017 R0
Status: Active

Instrument S/N 107676

Received From AST Cartwell

Date 06/24/2019

Analyst Dereck Walton

1. Reason for instrument return:

Preventive maintenance

2. Change instrument and case location in JusticeTrax Breath Alcohol Database

3. Ensure change in instrument status form is in JusticeTrax

4. Upload

5. Update agency information and polling status in DM Host Calling Program

6. Check for VOCs

From Intoximeters

7. Intoximeters Repair Paperwork (attached)