



State of Alaska
 Department of Public Safety
Scientific Crime Detection Laboratory

4805 Dr. Martin Luther King Jr. Ave.
 Anchorage, Alaska 99507
 David Kanaris, Chief, Forensic Laboratories

Breath Alcohol Calibration Report
Calibration Item: DataMaster DMT
Serial Number: 100700

Report Issue Date: 07/14/2023
Date Received: 2023-06-20
Condition of Item: Needs Repair

Method Used: DataMaster DMT Calibration / Current Breath Alcohol Procedure Manual
Calibration Range: 0.020 g/210L - 0.300 g/210L

Start Date 06/20/2023 - As Found - **End Date 06/20/2023** As Found - Evaluation Analyst: Colleen O'Bryant
 As Found Conditions - Barometric Pressure (inHg): 29.80 Barometer Used: Fisher S/N 221250180 Temperature (C): 22.2
 Relative Humidity (%): 51

As-Found Data									
Nominal Values (g/210L)	Target Values (g/210L)	Lot Number	Results (g/210L)					Mean (g/210L)	
0.020	0.020	AG217902	0.020	0.021	0.020	0.020	0.020	0.020	0.020
0.080	0.080	09523080A1	0.080	0.080	0.080	0.080	0.080	0.080	0.080
0.300	0.299	AG217903	0.302	0.302	0.302	0.302	0.302	0.303	0.302

Recommend replacing touch screen.

Start Date 06/27/2023 - As Left - **End Date 06/27/2023**

As Left Conditions - Barometric Pressure (inHg): 29.61 Barometer Used: Fisher S/N 221250180 Temperature (C): 23.2 Relative Humidity (%): 40

As-Left Data										
Nominal Values (g/210L)	Target Values (g/210L)	Lot Number	Results (g/210L)					Mean (g/210L)	Uncertainty (g/210L)	
0.020	0.020	AG217902	0.020	0.020	0.020	0.020	0.020	0.020	+/- 0.002	
0.040	0.040	AG217904	0.040	0.040	0.040	0.040	0.040	0.040	+/- 0.003	
0.080	0.079	09523080A1	0.080	0.080	0.079	0.080	0.080	0.080	+/- 0.004	
0.100	0.099	AG211003	0.100	0.100	0.100	0.099	0.100	0.100	+/- 0.004	
0.150	0.148	AG134205	0.150	0.149	0.150	0.149	0.150	0.150	+/- 0.005	
0.300	0.297	AG217903	0.300	0.300	0.300	0.300	0.300	0.300	+/- 0.011	

Calibration is performed using dry gas standards traceable to NIST. Standards are sourced from an external provider accredited to ISO/IEC 17025 and ISO 17034 standards. Internal instrument barometer readings are traceable to NIST. External barometers are calibrated by an external calibration provider accredited to the ISO/IEC 17025 standard.

The As-Left results listed above include the estimated measurement uncertainty with a coverage factor k=2 and an expanded coverage probability of 95%. The estimated uncertainty applies to the calibration process and not to subject breath samples.

This report applies only to the item being calibrated and shall not be reproduced, except in full, without written approval from the laboratory.

Results authorized by Colleen O'Bryant on 07/06/2023

colleen.obryant@alaska.gov

Technically and administratively reviewed by Brandi Barnett on 07/14/2023

Alaska Scientific Crime Detection Laboratory

DataMaster Calibration Form

Effective: 10/10/2022

Version: 3.0

Instrument #: 100700 Analyst: Tatiana Malnoske Start Date: 06/27/2023

- Diagnostic Test (attached)
- Adjust voltages, if needed
 - Lamp (2.60 V Maximum)
 - Bias
 - Cooler (2.00 V Maximum)
- Internal Standard value \leq 3.00%
 - Chopper Frequency
 - Detector Voltage (-0.100V to 0.100V)
- Set RFI Sensitivity
- Adjust internal barometer reading to match NIST-traceable barometer
- Adjust volume setting, if needed (1.50 L +/- 0.10 L)
- Print Technician Screen (attached)
- Print Calibration Factors (attached)
- Minimum Volume Test 1.0 L (1.3L to 1.7L)

Data Interpretation

06/27/2023

2023.07.06

TLM

COB

As Left Conditions:

Barometric Pressure (in Hg): 29.01 Temperature (C): 23.2 Relative Humidity (%): 40%

Linearity Test (attached)

0.020g/210L Mnf:	<u>Airgas</u>	Lot:	<u>AG217902</u>	Exp:	<u>06/28/2024</u>
0.040g/210L Mnf:	<u>Airgas</u>	Lot:	<u>AG217904</u>	Exp:	<u>06/28/2024</u>
0.080g/210L Mnf:	<u>ILMO</u>	Lot:	<u>0AS23080A1</u>	Exp:	<u>05/05/2025</u>
0.100g/210L Mnf:	<u>Airgas</u>	Lot:	<u>AG211003</u>	Exp:	<u>04/20/2024</u>
0.150g/210L Mnf:	<u>Airgas</u>	Lot:	<u>AG134205</u>	Exp:	<u>12/08/2023</u>
0.300g/210L Mnf:	<u>Airgas</u>	Lot:	<u>AG217903</u>	Exp:	<u>06/28/2024</u>

Functional Tests (attached)

- Incomplete
- Maximum Attempts Exceeded
- Interference (Volatiles)
- Invalid
- Radio Frequency Interference (RFI)

Acetone Mnf: Lifeloc Lot: 22HP41201 Level: 0.100 Exp: 08/24/2024

Simulator S/N: DK4176 Verification Performed?

Isopropanol Mnf: Lifeloc Lot: 22HP41201 Level: 0.100 Exp: 08/24/2024

Simulator S/N: DK4177 Verification Performed?

Neg. Control Mnf: Lifeloc Lot: 21KP41172 Level: 0.100 Exp: 11/09/2023

Simulator S/N: DK6267 Verification Performed?

Calibrated External Thermometer:

Thermometer S/N: 210A58384 Calibration Due Date: 11/09/2023

- Verify stylus is calibrated
 - Reset Options
 - Clear Memory
 - Verification of Calibration (VOC)
 - Non-Drinking Subject Test
 - Upload
 - Remove Tank Prior to Shutdown
- Calibration End Date: 06/27/2023

DIAGNOSTIC RECORD

Alaska Department of Public Safety
DATAMASTER dmt: 100700

Date: 06/27/2023
Time: 07:55:16

VERSIONS
DMT: 3.02
PIC: 3.03
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 49.0°C PASSED
(44.0°C - 52.0°C)
Breath Tube = 48.1°C PASSED
(38.0°C - 50.0°C)

SETTINGS

Lamp Voltage = 1.70 V PASSED
(1.02 V - 2.38 V)
Cooler Voltage = 1.68 V PASSED
(0.98 V - 2.28 V)
Bias Voltage = 80 V PASSED
(49 V - 115 V)
Chopper Freq = 534 Hz PASSED
(475 Hz - 575 Hz)
Barometer = 29.6 in

PUMP INFO

Flow Rate = 4.646 L/M PASSED
(3.500 L/M - 6.500 L/M)

DETECTOR INFO

PUMP ON PASSED
(0.002 V <= 0.010 V)
PUMP OFF PASSED
(0.002 V <= 0.010 V)

FILTER INFO

Filter 1 PASSED
Filter 2 PASSED
Filter 3 PASSED

INTERNAL STANDARD PASSED

Xq = 0.080 0.97% (0.00% - 4.00%)

Technician Screen

DATAMASTER dmt: 100700

Date: 06/27/2023
Time: 8:04 AM

Temperatures (°C)
Sample Cell: 48.7°C
Breath Tube: 47.6°C

Current Barometer 29.6 in
Volume (Ltr) 1.41

Settings
Lamp: 1.81 V
Bias: 79 V
Cooler: 1.53 V
Chopper: 532 Hz

Voltages (V)
Flow: 1.07 V
Detector: 0.084 V

CALIBRATION FACTORS

Alaska Department of Public Safety
DATAMASTER dmt: 100700

Date: 11/02/2011
Time: 08:45:04

Ca = 0.1000
CAL = 0.979351 0.800 <= CAL < 1.200
b1 = 0.0004 0.0000 <= b1 < 0.0040
b2 = 0.0040 0.0010 <= b2 < 0.0100
b3 = 0.0001 0.0000 <= b3 < 0.0040
Xq = 0.0805 0.0500 <= Xq < 0.2000
a21 = 1.149740 1.050 <= a21 < 1.300
a31 = 0.537738 0.300 <= a31 < 0.800

SUBJECT TEST

Alaska Department of Public Safety
DATAMASTER dmt: 100700

Date: 06/27/2023
Time: 08:21:54

SUBJECT NAME:
TEST TEST
LICENSE #: XXX


OPERATOR NAME:
TATIANA MALNOSKE
OPERATOR'S #: 10437

DEPT/AGENCY:
MINIMUM VOLUME TEST
CASE/REPORT: XXX
TEST TYPE: TEST

EXTERNAL STANDARD INFORMATION
NOMINAL: 0.080
TARGET AT 29.61: 0.079
LOT #: 09523080A1
EXPIRATION: 05/05/2025

DIAGNOSTIC CHECK	PASSED	08:22
BLANK TEST	0.000	08:23
INTERNAL STANDARD	VERIFIED	08:23
EXTERNAL STANDARD	0.079	08:23
BLANK TEST	0.000	08:24
SUBJECT SAMPLE (Vol=1.58L)	0.000	08:25
BLANK TEST	0.000	08:26
EXTERNAL STANDARD	0.080	08:26
BLANK TEST	0.000	08:27

TML



LINEARITY CHECK

Alaska Department of Public Safety
DATAMASTER dmt: 100700


Date: 06/27/2023
Time: 08:30:50

SUPERVISOR NAME:
TATIANA MALNOSKE
OPERATOR'S #: 10437

TARGET	AVERAGE	STD. DEV.
0.020	0.0200	0.0000
0.040	0.0400	0.0000
0.079	0.0798	0.0004
0.099	0.0998	0.0004
0.148	0.1496	0.0005
0.297	0.3000	0.0000

SLOPE = 1.011198
R² = 0.999998

TJM



LINEARITY CHECK: Control Test 1

 Alaska Department of Public Safety
 DATAMASTER dmt: 100700






Date: 06/27/2023
 Time: 08:30:52

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.020
 TARGET AT 29.60: 0.020
 LOT #: AG217902
 EXPIRATION: 06/28/2024
 TANK PRESSURE: 797 psi

BLANK TEST	0.000	08:32
INTERNAL STANDARD	VERIFIED	08:32
EXTERNAL STANDARD	0.020	08:33
BLANK TEST	0.000	08:34
EXTERNAL STANDARD	0.020	08:34
BLANK TEST	0.000	08:35
EXTERNAL STANDARD	0.020	08:35
BLANK TEST	0.000	08:36
EXTERNAL STANDARD	0.020	08:36
BLANK TEST	0.000	08:37
EXTERNAL STANDARD	0.020	08:37
BLANK TEST	0.000	08:38

Average = 0.0200
 Std Dev = 0.0000



LINEARITY CHECK: Control Test 2

 Alaska Department of Public Safety
 DATAMASTER dmt: 100700






Date: 06/27/2023
 Time: 08:40:56

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.040
 TARGET AT 29.61: 0.040
 LOT #: AG217904
 EXPIRATION: 06/28/2024
 TANK PRESSURE: 1078 psi

BLANK TEST	0.000	08:42
INTERNAL STANDARD	VERIFIED	08:42
EXTERNAL STANDARD	0.040	08:42
BLANK TEST	0.000	08:43
EXTERNAL STANDARD	0.040	08:43
BLANK TEST	0.000	08:44
EXTERNAL STANDARD	0.040	08:44
BLANK TEST	0.000	08:45
EXTERNAL STANDARD	0.040	08:45
BLANK TEST	0.000	08:46
EXTERNAL STANDARD	0.040	08:46
BLANK TEST	0.000	08:47

Average = 0.0400
 Std Dev = 0.0000



LINEARITY CHECK: Control Test 3

 Alaska Department of Public Safety
 DATAMASTER dmt: 100700






Date: 06/27/2023
 Time: 08:49:50

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
 TARGET AT 29.62: 0.079
 LOT #: 09523080A1
 EXPIRATION: 05/05/2025
 TANK PRESSURE: 770 psi

BLANK TEST	0.000	08:51
INTERNAL STANDARD	VERIFIED	08:51
EXTERNAL STANDARD	0.080	08:51
BLANK TEST	0.000	08:52
EXTERNAL STANDARD	0.080	08:52
BLANK TEST	0.000	08:53
EXTERNAL STANDARD	0.079	08:53
BLANK TEST	0.000	08:54
EXTERNAL STANDARD	0.080	08:54
BLANK TEST	0.000	08:55
EXTERNAL STANDARD	0.080	08:55
BLANK TEST	0.000	08:56

Average = 0.0798
 Std Dev = 0.0004



LINEARITY CHECK: Control Test 4

Alaska Department of Public Safety
 DATAMASTER dmt: 100700






Date: 06/27/2023
 Time: 08:58:59

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.100
 TARGET AT 29.62: 0.099
 LOT #: AG211003
 EXPIRATION: 04/20/2024
 TANK PRESSURE: 615 psi

BLANK TEST	0.000	09:00
INTERNAL STANDARD	VERIFIED	09:00
EXTERNAL STANDARD	0.100	09:00
BLANK TEST	0.000	09:01
EXTERNAL STANDARD	0.100	09:01
BLANK TEST	0.000	09:02
EXTERNAL STANDARD	0.100	09:03
BLANK TEST	0.000	09:03
EXTERNAL STANDARD	0.099	09:04
BLANK TEST	0.000	09:05
EXTERNAL STANDARD	0.100	09:05
BLANK TEST	0.000	09:06

Average = 0.0998
 Std Dev = 0.0004



LINEARITY CHECK: Control Test 5

Alaska Department of Public Safety
 DATAMASTER dmt: 100700






Date: 06/27/2023
 Time: 09:07:37

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.150
 TARGET AT 29.61: 0.148
 LOT #: AG134205
 EXPIRATION: 12/08/2023
 TANK PRESSURE: 739 psi

BLANK TEST	0.000	09:08
INTERNAL STANDARD	VERIFIED	09:09
EXTERNAL STANDARD	0.150	09:09
BLANK TEST	0.000	09:10
EXTERNAL STANDARD	0.149	09:10
BLANK TEST	0.000	09:11
EXTERNAL STANDARD	0.150	09:11
BLANK TEST	0.000	09:12
EXTERNAL STANDARD	0.149	09:12
BLANK TEST	0.000	09:13
EXTERNAL STANDARD	0.150	09:13
BLANK TEST	0.000	09:14

Average = 0.1496
 Std Dev = 0.0005



LINEARITY CHECK: Control Test 6

Alaska Department of Public Safety
 DATAMASTER dmt: 100700






Date: 06/27/2023
 Time: 09:23:19

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.300
 TARGET AT 29.62: 0.297
 LOT #: AG217903
 EXPIRATION: 06/28/2024
 TANK PRESSURE: 226 psi

BLANK TEST	0.000	09:24
INTERNAL STANDARD	VERIFIED	09:24
EXTERNAL STANDARD	0.300	09:24
BLANK TEST	0.000	09:25
EXTERNAL STANDARD	0.300	09:26
BLANK TEST	0.000	09:27
EXTERNAL STANDARD	0.300	09:27
BLANK TEST	0.000	09:28
EXTERNAL STANDARD	0.300	09:28
BLANK TEST	0.000	09:29
EXTERNAL STANDARD	0.300	09:29
BLANK TEST	0.000	09:30

Average = 0.3000
 Std Dev = 0.0000



SUBJECT TEST

Alaska Department of Public Safety
DATAMASTER dmt: 100700

Date: 06/27/2023
Time: 10:00:40

SUBJECT NAME:
TEST TEST
LICENSE #: XXX


OPERATOR NAME:
TATIANA MALNOSKE
OPERATOR'S #: 10437

DEPT/AGENCY:
INCOMPLETE
CASE/REPORT: XXX
TEST TYPE: TEST

EXTERNAL STANDARD INFORMATION
NOMINAL: 0.080
TARGET AT 29.62: 0.079
LOT #: 09523080A1
EXPIRATION: 05/05/2025

DIAGNOSTIC CHECK	PASSED	10:02
BLANK TEST	0.000	10:02
INTERNAL STANDARD	VERIFIED	10:02
EXTERNAL STANDARD	0.080	10:02
BLANK TEST	0.000	10:03
SUBJECT SAMPLE	INCOMPLETE	10:05

TM



SUBJECT TEST

Alaska Department of Public Safety
DATAMASTER dmt: 100700

Date: 06/27/2023
Time: 10:10:16

SUBJECT NAME:
TEST TEST
LICENSE #: XXX

OPERATOR NAME:
TATIANA MALNOSKE
OPERATOR'S #: 10437


DEPT/AGENCY:
MAXIMUM ATTEMPTS EXCEEDED
CASE/REPORT: XXX
TEST TYPE: TEST

EXTERNAL STANDARD INFORMATION
NOMINAL: 0.080
TARGET AT 29.62: 0.079
LOT #: 09523080A1
EXPIRATION: 05/05/2025

DIAGNOSTIC CHECK	PASSED	10:11
BLANK TEST	0.000	10:11
INTERNAL STANDARD	VERIFIED	10:11
EXTERNAL STANDARD	0.080	10:11
BLANK TEST	0.000	10:12
SUBJECT SAMPLE	XXX	10:13

Maximum Attempts Exceeded

TMM



SUBJECT TEST

Alaska Department of Public Safety
DATAMASTER dmt: 100700

Date: 06/27/2023
Time: 10:33:35

SUBJECT NAME:
TEST TEST
LICENSE #: XXX


OPERATOR NAME:
TATIANA MALNOSKE
OPERATOR'S #: 10437

DEPT/AGENCY:
INVALID
CASE/REPORT: XXX
TEST TYPE: TEST

EXTERNAL STANDARD INFORMATION
NOMINAL: 0.080
TARGET AT 29.62: 0.079
LOT #: 09523080A1
EXPIRATION: 05/05/2025

DIAGNOSTIC CHECK	PASSED	10:34
BLANK TEST	0.000	10:34
INTERNAL STANDARD	VERIFIED	10:34
EXTERNAL STANDARD	0.079	10:35
BLANK TEST	0.000	10:35
SUBJECT SAMPLE	INVALID	10:36

TAM



SUBJECT TEST

Alaska Department of Public Safety
DATAMASTER dmt: 100700

Date: 06/27/2023
Time: 11:16:48

SUBJECT NAME:
TEST TEST
LICENSE #: XXX

OPERATOR NAME:
TATIANA MALNOSKE
OPERATOR'S #: 10437


DEPT/AGENCY:
RADIO FREQUENCY INTERFERENCE (RFI)
CASE/REPORT: XXX
TEST TYPE: TEST

EXTERNAL STANDARD INFORMATION
NOMINAL: 0.080
TARGET AT 29.62: 0.079
LOT #: 09523080A1
EXPIRATION: 05/05/2025

DIAGNOSTIC CHECK	PASSED	11:17
BLANK TEST	0.000	11:18
INTERNAL STANDARD	VERIFIED	11:18
EXTERNAL STANDARD	0.080	11:18
BLANK TEST	0.000	11:19

RFI DETECTED

RM



SUBJECT TEST

Alaska Department of Public Safety
DATAMASTER dmt: 100700

Date: 06/27/2023
Time: 11:10:17

SUBJECT NAME:
TEST TEST
LICENSE #: XXX

OPERATOR NAME:
TATIANA MALNOSKE
OPERATOR'S #: 10437

DEPT/AGENCY:
INTERFERENCE DETECTED - ACETONE
CASE/REPORT: XXX
TEST TYPE: TEST


EXTERNAL STANDARD INFORMATION
NOMINAL: 0.080
TARGET AT 29.62: 0.079
LOT #: 09523080A1
EXPIRATION: 05/05/2025

DIAGNOSTIC CHECK	PASSED	11:11
BLANK TEST	0.000	11:11
INTERNAL STANDARD	VERIFIED	11:11
Xq = 0.081 (0.22%)		
EXTERNAL STANDARD	0.080	11:11
X[1] = 0.0797 (0.0004) (0.0020)		
BLANK TEST	0.000	11:12
SUBJECT SAMPLE	XXX	11:13
X[1] = 0.1087 (-0.0178) (0.0013)		

Interference Detected

Filter 2 Average : 0.0005
Filter 3 Average : 0.0022
TJM 06/27/2023

TJM



SUBJECT TEST

Alaska Department of Public Safety
DATAMASTER dmt: 100700

Date: 06/27/2023
Time: 10:56:34

SUBJECT NAME:
TEST TEST
LICENSE #: XXX

OPERATOR NAME:
TATIANA MALNOSKE
OPERATOR'S #: 10437


DEPT/AGENCY:
INTERFERENCE DETECTED - ISOPROPRANOL
CASE/REPORT: XXX
TEST TYPE: TEST

EXTERNAL STANDARD INFORMATION
NOMINAL: 0.080
TARGET AT 29.62: 0.079
LOT #: 09523080A1
EXPIRATION: 05/05/2025

DIAGNOSTIC CHECK	PASSED	10:57
BLANK TEST	0.000	10:57
INTERNAL STANDARD	VERIFIED	10:57
Xq = 0.080 (0.32%)		
EXTERNAL STANDARD	0.079	10:58
X[1] = 0.0793 (0.0005) (0.0028)		
BLANK TEST	0.000	10:58
SUBJECT SAMPLE	XXX	10:59
X[1] = 0.1244 (-0.0285) (0.0016)		

Interference Detected

TMM



SUBJECT TEST

Alaska Department of Public Safety
DATAMASTER dmt: 100700

Date: 06/27/2023
Time: 11:01:56

SUBJECT NAME:
TEST TEST
LICENSE #: XXX


OPERATOR NAME:
TATIANA MALNOSKE
OPERATOR'S #: 10437

DEPT/AGENCY:
INTERFERENCE DETECTED - NEGATIVE CONTROL
CASE/REPORT: XXX
TEST TYPE: TEST

EXTERNAL STANDARD INFORMATION
NOMINAL: 0.080
TARGET AT 29.62: 0.079
LOT #: 09523080A1
EXPIRATION: 05/05/2025

DIAGNOSTIC CHECK	PASSED	11:03
BLANK TEST	0.000	11:03
INTERNAL STANDARD	VERIFIED	11:03
Xq = 0.080 (0.54%)		
EXTERNAL STANDARD	0.080	11:03
X[1] = 0.0796 (0.0001) (0.0018)		
BLANK TEST	0.000	11:04
SUBJECT SAMPLE (Vol=1.98L)	0.095	11:05
X[1] = 0.0952 (0.0005) (0.0005)		
BLANK TEST	0.000	11:06
EXTERNAL STANDARD	0.079	11:06
X[1] = 0.0794 (0.0010) (0.0022)		
BLANK TEST	0.000	11:07

TML



SUBJECT TEST

Alaska Department of Public Safety
DATAMASTER dmt: 100700

Date: 06/27/2023
Time: 12:19:08

SUBJECT NAME:
TEST TEST
LICENSE #: XXX

OPERATOR NAME:
TATIANA MALNOSKE
OPERATOR'S #: 10437

DEPT/AGENCY:
NON-DRINKING SUBJECT TEST
CASE/REPORT: XXX
TEST TYPE: TEST

EXTERNAL STANDARD INFORMATION
NOMINAL: 0.080
TARGET AT 29.62: 0.079
LOT #: 09523080A1
EXPIRATION: 05/05/2025

DIAGNOSTIC CHECK	PASSED	12:20
BLANK TEST	0.000	12:20
INTERNAL STANDARD	VERIFIED	12:20
EXTERNAL STANDARD	0.080	12:20
BLANK TEST	0.000	12:21
SUBJECT SAMPLE (Vol=1.67L)	0.000	12:21
BLANK TEST	0.000	12:22
EXTERNAL STANDARD	0.080	12:22
BLANK TEST	0.000	12:23

TML

