

Alaska Scientific Crime Detection Laboratory

DataMaster Check In Form

Issued: 7/10/2015
Effective: 7/10/2015

Version: DMCIF 2015 R1
Status: Active

Instrument S/N 100389

Received From AST/Gurwood

Date 10/5/16

Analyst Colleen O'Bryen

1. Reason for instrument return:

removed due to part closure.

2. Change instrument and case location in JusticeTrax Breath Alcohol Database

3. Ensure change in instrument status form is in JusticeTrax

4. Upload

5. Update agency information and polling status in DM Host Calling Program

6. Check for VOCs

From NPAS

7. NPAS Repair Paperwork (attached)